



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113792

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/02/2021	08:30
2	CAL2	1	Water	NA	NA	20.1	7.00	04/02/2021	08:31
3	CAL3	1	Water	NA	NA	20.1	10.03	04/02/2021	08:32
4	ICV	1	Water	NA	NA	20.2	6.99	04/02/2021	08:35
5	CCV1	1	Water	NA	NA	20.1	2.01	04/02/2021	08:36
6	M1898-01	1	Water	NA	NA	20.5	7.25	04/02/2021	08:45
7	M1898-03	1	Water	NA	NA	20.5	5.20	04/02/2021	08:46
8	M1898-05	1	Water	NA	NA	20.6	5.33	04/02/2021	08:47
9	M1903-01	1	Water	NA	NA	20.5	6.67	04/02/2021	08:49
10	M1903-02	1	Water	NA	NA	20.7	4.19	04/02/2021	08:50
11	M1903-03	1	Water	NA	NA	20.6	4.53	04/02/2021	08:52
12	M1903-04	1	Water	NA	NA	20.7	6.50	04/02/2021	09:00
13	M1903-05	1	Water	NA	NA	20.8	6.24	04/02/2021	09:02
14	M1903-06	1	Water	NA	NA	20.7	6.50	04/02/2021	09:05
15	M1903-07	1	Water	NA	NA	20.7	6.83	04/02/2021	09:06
16	CCV2	1	Water	NA	NA	20.3	12.02	04/02/2021	09:10
17	M1903-08	1	Water	NA	NA	20.9	9.05	04/02/2021	09:12
18	M1903-09	1	Water	NA	NA	20.8	7.03	04/02/2021	09:15
19	M1903-09DUP	1	Water	NA	NA	20.9	7.04	04/02/2021	09:16
20	CCV3	1	Water	NA	NA	20.3	2.01	04/02/2021	09:19

WORKLIST(Hardcopy Internal Chain)

18 113792

WorkList Name : ph m1989

WorkList ID : 147675

Department : Wet-Chemistry

Date : 04-02-2021 08:20:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/01/2021	Water	M1898-01	✓ pH	Cool 4 deg C	PSEG03	O11	32998	04/01/2021	9040C
04/01/2021	Water	M1898-03	✓ pH	Cool 4 deg C	PSEG03	O11	32999	04/01/2021	9040C
04/01/2021	Water	M1898-05	✓ pH	Cool 4 deg C	PSEG03	O11	33000	04/01/2021	9040C
04/01/2021	Water	M1903-01	✓ pH	Cool 4 deg C	PSEG03	O12	30776	04/01/2021	9040C
04/01/2021	Water	M1903-02	✓ pH	Cool 4 deg C	PSEG03	O12	30791	04/01/2021	9040C
04/01/2021	Water	M1903-03	✓ pH	Cool 4 deg C	PSEG03	O12	30792	04/01/2021	9040C
04/01/2021	Water	M1903-04	✓ pH	Cool 4 deg C	PSEG03	O12	41726	04/01/2021	9040C
04/01/2021	Water	M1903-05	✓ pH	Cool 4 deg C	PSEG03	O12	30809	04/01/2021	9040C
04/01/2021	Water	M1903-06	✓ pH	Cool 4 deg C	PSEG03	O12	41749	04/01/2021	9040C
04/01/2021	Water	M1903-07	✓ pH	Cool 4 deg C	PSEG03	O12	31029	04/01/2021	9040C
04/01/2021	Water	M1903-08	✓ pH	Cool 4 deg C	PSEG03	O12	31030	04/01/2021	9040C
04/01/2021	Water	M1903-09	✓ pH	Cool 4 deg C	PSEG03	O12	31031	04/01/2021	9040C

Date/Time 04/02/21 08:25
 Received by: JP (Wet Chem)
 Relinquished by: KS (Wet Chem)

Date/Time 04/02/21 11:09
 Received by: KS (Wet Chem)
 Relinquished by: JP (Wet Chem)