



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113794

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	04/02/2021	09:40
2	M1898-01		1	10	100.00	>212.0	>212.0	04/02/2021	10:10
3	M1898-01DUP		1	10	100.00	>212.0	>212.0	04/02/2021	10:35
4	M1898-03		1	15	100.00	>212.0	>212.0	04/02/2021	10:58
5	M1898-05		1	12	100.00	>212.0	>212.0	04/02/2021	11:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB113794

WorkList Name : fp4221 WorkList ID : 147693 Department : Wet-Chemistry Date : 04-02-2021 09:13:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/06/2021	Water	M1898-01	Flash Point	Cool 4 deg C	PSEG03	O11	32998	04/01/2021	1010A
04/06/2021	Water	M1898-03	Flash Point	Cool 4 deg C	PSEG03	O11	32999	04/01/2021	1010A
04/06/2021	Water	M1898-05	Flash Point	Cool 4 deg C	PSEG03	O11	33000	04/01/2021	1010A

Date/Time 04/02/21 11:30
 Received by: SP 1624-06(m)
 Relinquished by: KS (watching)

Date/Time 04/02/21 09:15
 Received by: KS (watching)
 Relinquished by: SP (watching)