

LB113795


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38109



A= 0.0000e+000

B= 2.6187e-004

C= 3.3493e-002

Rho= 0.9998890

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.049	0.049	58	0.000	58					15
0.2	0.200	0.230	0.030	752	0.0 %	752					-5
2.5	2.500	2.381	-0.119	8965	0.0 %	8965					0
5.0	5.000	5.023	0.023	19054	0.0 %	19054					0
7.5	7.500	7.503	0.003	28524	0.0 %	28524					0
10.0	10.000	10.013	0.013	38109	0.0 %	38109					0

LB113795 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Apr 2021 11:54:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	58	-	0.0000	7470A	PPB	02 Apr 2021 11:55:05	S	Std
0.2 - 1	50.2	752	-	0.2000	7470A	PPB	02 Apr 2021 11:57:22	S	Std
2.5 - 1	52.5	8965	-	2.5000	7470A	PPB	02 Apr 2021 11:59:39	S	Std
5.0 - 1	55	19054	-	5.0000	7470A	PPB	02 Apr 2021 12:01:56	S	Std
7.5 - 1	57.5	28524	-	7.5000	7470A	PPB	02 Apr 2021 12:04:12	S	Std
10.0 - 1	51.0	38109	-	10.0000	7470A	PPB	02 Apr 2021 12:06:29	S	Std
ICV65 - 1	ICV65	13991	3.6974	-7470A	PPB	PPB	02 Apr 2021 12:09:15	U	SMPL
ICB65 - 1	ICB65	-336	-0.0545	-7470A	PPB	PPB	02 Apr 2021 12:11:29	U	SMPL
CCV24 - 1	CCV24	19487	5.1366	-7470A	PPB	PPB	02 Apr 2021 12:13:45	U	SMPL
CCB24 - 1	CCB24	-321	-0.0506	-7470A	PPB	PPB	02 Apr 2021 12:16:00	U	SMPL
CRA - 1	CRA	669	0.2087	-7470A	PPB	PPB	02 Apr 2021 12:18:15	U	SMPL
HighStd - 1	HighStd	37606	9.8815	-7470A	PPB	PPB	02 Apr 2021 12:20:31	U	SMPL
ChkStd - 1	ChkStd	25331	6.6670	-7470A	PPB	PPB	02 Apr 2021 12:22:46	U	SMPL
PB135495BL - 1	PBW	-360	-0.0608	-7470A	PPB	PPB	02 Apr 2021 12:25:03	U	SMPL
PB135495BS - 1	LCSW	15398	4.0658	-7470A	PPB	PPB	02 Apr 2021 12:27:21	U	SMPL
M1883-02 - 1	IEC-R2	-192	-0.0168	-7470A	PPB	PPB	02 Apr 2021 12:29:37	U	SMPL
M1883-04 - 1	V-211	-10	0.0309	-7470A	PPB	PPB	02 Apr 2021 12:31:54	U	SMPL
M1883-04DUP - 1	V-211DUP	16	0.0377	-7470A	PPB	PPB	02 Apr 2021 12:34:10	U	SMPL
M1883-04MS - 1	V-211MS	15628	4.1261	-7470A	PPB	PPB	02 Apr 2021 12:36:27	U	SMPL
M1883-04MSD - 1	V-211MSD	15630	4.1266	-7470A	PPB	PPB	02 Apr 2021 12:38:42	U	SMPL
CCV25 - 1	CCV25	19257	5.0764	-7470A	PPB	PPB	02 Apr 2021 12:40:58	U	SMPL
CCB25 - 1	CCB25	-337	-0.0548	-7470A	PPB	PPB	02 Apr 2021 12:43:14	U	SMPL
PB135495TB - 1	PB135473TB	-60	0.0178	-7470A	PPB	PPB	02 Apr 2021 12:45:31	U	SMPL
M1883-04LX5 - 1		-22	0.0277	-7470A	PPB	PPB	02 Apr 2021 12:47:46	U	SMPL
M1883-04A - 1		15554	4.1067	-7470A	PPB	PPB	02 Apr 2021 12:50:01	U	SMPL
CCV26 - 1	CCV26	19136	5.0447	-7470A	PPB	PPB	02 Apr 2021 12:52:16	U	SMPL
CCB26 - 1	CCB26	-330	-0.0529	-7470A	PPB	PPB	02 Apr 2021 12:54:34	U	SMPL