

LB113803


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

58208



A= 0.0000e+000

B= 1.7063e-004

C= 2.1336e-002

Rho= 0.9999138

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.023	0.023	11	0.000	11					
0.2	0.200	0.209	0.009	1097	0.0 %	1097					4
2.5	2.500	2.431	-0.069	14123	0.0 %	14123					-3
5.0	5.000	5.002	0.002	29191	0.0 %	29191					0
7.5	7.500	7.581	0.081	44304	0.0 %	44304					1
10.0	10.000	9.954	-0.046	58208	0.0 %	58208					0

LB113803 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 02 Apr 2021 14:05:30					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	58.2	11	-	0.0000	7471B	PPB	02 Apr 2021 14:08:33	S	Std
0.2 - 1	58.2	1097	-	0.2000	7471B	PPB	02 Apr 2021 14:10:50	S	Std
2.5 - 1	52.5	14123	-	2.5000	7471B	PPB	02 Apr 2021 14:13:07	S	Std
5.0 - 1	55	29191	-	5.0000	7471B	PPB	02 Apr 2021 14:15:24	S	Std
7.5 - 1	57.5	44304	-	7.5000	7471B	PPB	02 Apr 2021 14:17:40	S	Std
10.0 - 1	58	58208	-	10.0000	7471B	PPB	02 Apr 2021 14:19:57	S	Std
ICV66 - 1	ICV66	22169	3.8041	-	7471B	PPB	02 Apr 2021 14:22:43	U	SMPL
ICB66 - 1	ICB66	-271	-0.0249	-	7471B	PPB	02 Apr 2021 14:24:58	U	SMPL
CCV27 - 1	CCV27	28519	4.8877	-	7471B	PPB	02 Apr 2021 14:27:14	U	SMPL
CCB27 - 1	CCB27	-300	-0.0299	-	7471B	PPB	02 Apr 2021 14:29:29	U	SMPL
CRA - 1	CRA	1057	0.2017	-	7471B	PPB	02 Apr 2021 14:31:45	U	SMPL
HighStd - 1	HighStd	58215	9.9548	-	7471B	PPB	02 Apr 2021 14:34:00	U	SMPL
ChkStd - 1	ChkStd	41546	7.1105	-	7471B	PPB	02 Apr 2021 14:36:15	U	SMPL
PB135505BL - 1	PBS	-313	-0.0321	-	7471B	PPB	02 Apr 2021 14:38:33	U	SMPL
PB135505BS - 1	LCSS	21058	3.6146	-	7471B	PPB	02 Apr 2021 14:40:51	U	SMPL
M1854-01 - 1	TP-1S-(0-4)	14074	2.4228	-	7471B	PPB	02 Apr 2021 14:52:52	U	SMPL
M1854-01DUP - 1	TP-1S-(0-4)DUP	12320	2.1236	-	7471B	PPB	02 Apr 2021 14:55:08	U	SMPL
M1854-01MS - 1	TP-1S-(0-4)MS	37994	6.5044	-	7471B	PPB	02 Apr 2021 15:00:12	U	SMPL
M1854-01MSD - 1	TP-1S-(0-4)MSD	37840	6.4781	-	7471B	PPB	02 Apr 2021 15:02:28	U	SMPL
M1854-05 - 1	TP-1D-(4-10)	6450	1.1219	-	7471B	PPB	02 Apr 2021 15:04:43	U	SMPL
CCV28 - 1	CCV28	28899	4.9525	-	7471B	PPB	02 Apr 2021 15:06:59	U	SMPL
CCB28 - 1	CCB28	-97	0.0048	-	7471B	PPB	02 Apr 2021 15:09:14	U	SMPL
M1883-01 - 1	IEC-R2	161172	27.5228	-	7471B	PPB	02 Apr 2021 15:11:29	U	SMPL
M1883-03 - 1	V-211	17402	2.9907	-	7471B	PPB	02 Apr 2021 15:13:44	U	SMPL
M1885-01 - 1	SG1-8015-8025	718	0.1439	-	7471B	PPB	02 Apr 2021 15:16:02	U	SMPL
M1885-02 - 1	SG1-8014-8024	499	0.1065	-	7471B	PPB	02 Apr 2021 15:18:18	U	SMPL
M1885-03 - 1	SG1-8013-8023	318	0.0756	-	7471B	PPB	02 Apr 2021 15:20:33	U	SMPL
M1885-04 - 1	SG1-8012-8022	293	0.0713	-	7471B	PPB	02 Apr 2021 15:22:49	U	SMPL
M1885-05 - 1	SG1-8011-8021	313	0.0747	-	7471B	PPB	02 Apr 2021 15:25:06	U	SMPL
M1885-06 - 1	SG2-8035-8045	285	0.0700	-	7471B	PPB	02 Apr 2021 15:27:22	U	SMPL
M1885-07 - 1	SG2-8034-8044	185	0.0529	-	7471B	PPB	02 Apr 2021 15:29:39	U	SMPL
M1885-08 - 1	SG2-8033-8043	205	0.0563	-	7471B	PPB	02 Apr 2021 15:31:55	U	SMPL
CCV29 - 1	CCV29	28979	4.9661	-	7471B	PPB	02 Apr 2021 15:34:10	U	SMPL
CCB29 - 1	CCB29	-32	0.0159	-	7471B	PPB	02 Apr 2021 15:36:26	U	SMPL
M1885-09 - 1	SG2-8032-8042	346	0.0804	-	7471B	PPB	02 Apr 2021 15:38:41	U	SMPL
M1885-10 - 1	EMPTY	365	0.0836	-	7471B	PPB	02 Apr 2021 15:40:57	U	SMPL
M1886-01 - 1	RO-124035	2457	0.4406	-	7471B	PPB	02 Apr 2021 15:43:12	U	SMPL
M1886-03 - 1	RO-124027	2906	0.5172	-	7471B	PPB	02 Apr 2021 15:45:28	U	SMPL
M1886-05 - 1	124018-A	2391	0.4293	-	7471B	PPB	02 Apr 2021 15:47:43	U	SMPL
M1886-07 - 1	RO-124037	3153	0.5593	-	7471B	PPB	02 Apr 2021 15:49:59	U	SMPL
M1887-01 - 1	TR-03-033121	1606	0.2954	-	7471B	PPB	02 Apr 2021 15:52:16	U	SMPL
PB135507BL - 1	PBS	126	0.0428	-	7471B	PPB	02 Apr 2021 15:54:33	U	SMPL
PB135507BS - 1	LCSS	23669	4.0601	-	7471B	PPB	02 Apr 2021 15:56:50	U	SMPL
M1888-01 - 1	TP-3S-(0-4)	99379	16.9788	-	7471B	PPB	02 Apr 2021 15:59:06	U	SMPL
CCV30 - 1	CCV30	29125	4.9911	-	7471B	PPB	02 Apr 2021 16:01:21	U	SMPL
CCB30 - 1	CCB30	-67	0.0099	-	7471B	PPB	02 Apr 2021 16:03:38	U	SMPL
M1888-04 - 1	TP-3D-(4-10)	6375	1.1091	-	7471B	PPB	02 Apr 2021 16:05:53	U	SMPL
M1889-01 - 1	TP-5-(0-4)	118218	20.1934	-	7471B	PPB	02 Apr 2021 16:08:08	U	SMPL
M1894-01 - 1	20210155-CP-COMPOSITE	51107	8.7419	-	7471B	PPB	02 Apr 2021 16:10:24	U	SMPL
M1894-02 - 1	20210156-AMENDED-CP-COMPOSITE	45851	7.8451	-	7471B	PPB	02 Apr 2021 16:12:41	U	SMPL
M1897-02 - 1	TP-1	1590	0.2926	-	7471B	PPB	02 Apr 2021 16:14:58	U	SMPL
M1897-02DUP - 1	TP-1DUP	1808	0.3298	-	7471B	PPB	02 Apr 2021 16:20:06	U	SMPL
M1897-02MS - 1	TP-1MS	24753	4.2450	-	7471B	PPB	02 Apr 2021 16:22:22	U	SMPL
M1897-02MSD - 1	TP-1MSD	24411	4.1867	-	7471B	PPB	02 Apr 2021 16:24:39	U	SMPL
M1897-04 - 1	TP-2	2080	0.3763	-	7471B	PPB	02 Apr 2021 16:26:56	U	SMPL
M1900-01 - 1	CHRT-26586	13366	2.3020	-	7471B	PPB	02 Apr 2021 16:29:11	U	SMPL
CCV31 - 1	CCV31	30121	5.1610	-	7471B	PPB	02 Apr 2021 16:31:26	U	SMPL
CCB31 - 1	CCB31	-18	0.0183	-	7471B	PPB	02 Apr 2021 16:33:42	U	SMPL
M1900-03 - 1	CHRT-20111	3992	0.7025	-	7471B	PPB	02 Apr 2021 16:35:58	U	SMPL
M1901-01 - 1	SP-15TH-STREET	91467	15.6287	-	7471B	PPB	02 Apr 2021 16:38:13	U	SMPL
M1908-01 - 1	SB-1(2-2.5)	2052	0.3715	-	7471B	PPB	02 Apr 2021 16:40:29	U	SMPL
M1908-02 - 1	SB-2(2-2.5)	6807	1.1828	-	7471B	PPB	02 Apr 2021 16:42:46	U	SMPL
M1908-05 - 1	SB-5(3.0-3.5)	807	0.1590	-	7471B	PPB	02 Apr 2021 16:45:02	U	SMPL
M1908-06 - 1	SB-6(3.0-3.5)	2920	0.5196	-	7471B	PPB	02 Apr 2021 16:47:18	U	SMPL
M1909-01 - 1	FS-SB-10-5-10	77481	13.2423	-	7471B	PPB	02 Apr 2021 16:49:34	U	SMPL
M1909-02 - 1	FS-SB-10-26-27	110815	18.9302	-	7471B	PPB	02 Apr 2021 16:51:51	U	SMPL
M1909-03 - 1	FS-SB-17-12-13	826	0.1623	-	7471B	PPB	02 Apr 2021 16:54:09	U	SMPL
M1911-01 - 1	TP-6S-(0-4)	762717	HIGH	-	7471B	PPB	02 Apr 2021 16:56:27	X	
CCV32 - 1	CCV32	30618	5.2458	-	7471B	PPB	02 Apr 2021 16:58:14	U	SMPL
CCB32 - 1	CCB32	-114	0.0019	-	7471B	PPB	02 Apr 2021 17:00:39	U	SMPL
M1911-04 - 1	TP-6D-(4-10)	577769	HIGH	-	7471B	PPB	02 Apr 2021 17:02:56	X	
M1854-01LX5 - 1		2813	0.5013	-	7471B	PPB	02 Apr 2021 17:04:44	U	SMPL
M1854-01A - 1		39184	6.7075	-	7471B	PPB	02 Apr 2021 17:07:02	U	SMPL
M1897-02LX5 - 1		264	0.0664	-	7471B	PPB	02 Apr 2021 17:09:18	U	SMPL
M1897-02A - 1		25038	4.2937	-	7471B	PPB	02 Apr 2021 17:11:35	U	SMPL
M1883-01DLX5 - 1		35157	6.0203	-	7471B	PPB	02 Apr 2021 17:13:51	U	SMPL
M1888-01DLX5 - 1		20170	3.4630	-	7471B	PPB	02 Apr 2021 17:16:07	U	SMPL
M1889-01DLX5 - 1		23856	4.0920	-	7471B	PPB	02 Apr 2021 17:18:24	U	SMPL
M1901-01DLX5 - 1		18125	3.1141	-	7471B	PPB	02 Apr 2021 17:20:41	U	SMPL
M1911-01DLX10 - 1		177864	30.3710	-	7471B	PPB	02 Apr 2021 17:22:57	U	SMPL
CCV33 - 1	CCV33	29558	5.0649	-	7471B	PPB	02 Apr 2021 17:25:13	U	SMPL
CCB33 - 1	CCB33	-88	0.0063	-	7471B	PPB	02 Apr 2021 17:27:32	U	SMPL
M1911-04DLX10 - 1		102946	17.5875	-	7471B	PPB	02 Apr 2021 17:29:48	U	SMPL
M1909-01DLX5 - 1		15167	2.6093	-	7471B	PPB	02 Apr 2021 17:32:04	U	SMPL
M1909-02DLX5 - 1		21976	3.7712	-	7471B	PPB	02 Apr 2021 17:34:22	U	SMPL
M1911-01DL2X100 - 1		19131	3.2857	-	7471B	PPB	02 Apr 2021 17:36:38	U	SMPL
M1911-04DL2X50 - 1		20795	3.5697	-	7471B	PPB	02 Apr 2021 17:38:54	U	SMPL
CCV34 - 1	CCV34	29886	5.1209	-	7471B	PPB	02 Apr 2021 17:41:10	U	SMPL
CCB34 - 1	CCB34	-42	0.0142	-	7471B	PPB	02 Apr 2021 17:43:27	U	SMPL