



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113814

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	04/02/2021	12:45
2	M1923-01		1	12	100.00	>212.0	>212.0	04/02/2021	13:15
3	M1923-01DUP		1	12	100.00	>212.0	>212.0	04/02/2021	13:40
4	M1923-03		1	10	100.00	>212.0	>212.0	04/02/2021	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB113814

WORKLIST(Hardcopy Internal Chain)

WorkList Name: FP4221-1 WorkList ID: 147697 Department: Wet-Chemistry Date: 04-02-2021 12:35:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/05/2021	Water	M1923-01	Flash Point	Cool 4 deg C	PSEG03	O13	WATER-CONTAINMENT-1	04/02/2021	1010A
04/05/2021	Water	M1923-03	Flash Point	Cool 4 deg C	PSEG03	O13	WATER-CONTAINMENT-2	04/02/2021	1010A

Date/Time 04/02/21 12:20
Received by: KS (Wetchem)
Relinquished by: RCCExt. (ab)

Date/Time 04/02/21 14:15
Received by: RCCExt. (ab)
Relinquished by: KS (Wetchem)