


LB113818

7470A

INSTRUMENT ID: CV1

 μ Abs.:

38134



A= 0.0000e+000

B= 2.6005e-004

C= 6.4772e-003

Rho= 0.9996014

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Y01
0.0	0.000	0.011	0.011	16	0.000	16					-
0.2	0.200	0.216	0.016	805	0.0 %	805					8
2.5	2.500	2.344	-0.156	8987	0.0 %	8987					-6
5.0	5.000	5.184	0.184	19910	0.0 %	19910					4
7.5	7.500	7.523	0.023	28902	0.0 %	28902					0
10.0	10.000	9.923	-0.077	38134	0.0 %	38134					-1

LB113818 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 05 Apr 2021 11:25:16							
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	16	-	0.0000	7470A	PPB	05 Apr 2021 11:38:26	S	Std
0.2 - 1	50.2	805	-	0.2000	7470A	PPB	05 Apr 2021 11:40:43	S	Std
2.5 - 1	52.5	8987	-	2.5000	7470A	PPB	05 Apr 2021 11:43:00	S	Std
5.0 - 1	55	19910	-	5.0000	7470A	PPB	05 Apr 2021 11:45:17	S	Std
7.5 - 1	57.5	28902	-	7.5000	7470A	PPB	05 Apr 2021 11:52:41	S	Std
10.0 - 1	59	38134	-	10.0000	7470A	PPB	05 Apr 2021 11:57:31	S	Std
ICV69 - 1	ICV69	15773	4.1083	-	7470A	PPB	05 Apr 2021 12:00:15	U	SMPL
ICB69 - 1	ICB69	76	0.0262	-	7470A	PPB	05 Apr 2021 12:11:50	U	SMPL
CCV35 - 1	CCV35	20431	5.3196	-	7470A	PPB	05 Apr 2021 12:14:05	U	SMPL
CCB35 - 1	CCB35	-194	-0.0440	-	7470A	PPB	05 Apr 2021 12:16:20	U	SMPL
CRA - 1	CRA	801	0.2148	-	7470A	PPB	05 Apr 2021 12:20:57	U	SMPL
HighStd - 1	HighStd	37693	9.8086	-	7470A	PPB	05 Apr 2021 12:23:11	U	SMPL
ChkStd - 1	ChkStd	25905	6.7431	-	7470A	PPB	05 Apr 2021 12:25:26	U	SMPL
PB135521BL - 1	PBW	-52	-0.0070	-	7470A	PPB	05 Apr 2021 12:30:19	U	SMPL
PB135521BS - 1	LCSW	15021	3.9127	-	7470A	PPB	05 Apr 2021 12:32:34	U	SMPL
M1894-04 - 1	0156-AMENDED-CP-COMP	-145	-0.0312	-	7470A	PPB	05 Apr 2021 12:34:50	U	SMPL
M1894-04DUP - 1	0156-AMENDED-CP-COMPDUP	11	0.0093	-	7470A	PPB	05 Apr 2021 12:37:07	U	SMPL
M1894-04MS - 1	0156-AMENDED-CP-COMPMS	15922	4.1470	-	7470A	PPB	05 Apr 2021 12:39:24	U	SMPL
M1894-04MSD - 1	0156-AMENDED-CP-COMPMSD	15918	4.1460	-	7470A	PPB	05 Apr 2021 12:41:40	U	SMPL
PB135522BL - 1	PBW	5	0.0078	-	7470A	PPB	05 Apr 2021 12:46:25	U	SMPL
CCV36 - 1	CCV36	20270	5.2777	-	7470A	PPB	05 Apr 2021 12:48:40	U	SMPL
CCB36 - 1	CCB36	-24	0.0002	-	7470A	PPB	05 Apr 2021 12:53:30	U	SMPL
PB135522BS - 1	LCSW	14917	3.8857	-	7470A	PPB	05 Apr 2021 12:55:45	U	SMPL
M1894-03 - 1	0154-FIELD-BLANK	-225	-0.0520	-	7470A	PPB	05 Apr 2021 12:58:00	U	SMPL
M1903-01 - 1	30776	11740	3.0595	-	7470A	PPB	05 Apr 2021 13:00:15	U	SMPL
M1910-01 - 1	FS-SB-17	737	0.1981	-	7470A	PPB	05 Apr 2021 13:02:30	U	SMPL
M1923-01 - 1	WATER-CONTAINMENT-1	162	0.0486	-	7470A	PPB	05 Apr 2021 13:04:47	U	SMPL
M1923-03 - 1	WATER-CONTAINMENT-2	-38	-0.0034	-	7470A	PPB	05 Apr 2021 13:07:03	U	SMPL
M1923-03DUP - 1	WATER-CONTAINMENT-2DUP	-14	0.0028	-	7470A	PPB	05 Apr 2021 13:09:20	U	SMPL
M1923-03MS - 1	WATER-CONTAINMENT-2MS	15006	3.9088	-	7470A	PPB	05 Apr 2021 13:11:37	U	SMPL
M1923-03MSD - 1	WATER-CONTAINMENT-2MSD	15075	3.9268	-	7470A	PPB	05 Apr 2021 13:13:53	U	SMPL
CCV37 - 1	CCV37	20263	5.2759	-	7470A	PPB	05 Apr 2021 13:16:09	U	SMPL
CCB37 - 1	CCB37	-35	-0.0026	-	7470A	PPB	05 Apr 2021 13:18:24	U	SMPL
PB135521TB - 1	PB135488TB	-59	-0.0089	-	7470A	PPB	05 Apr 2021 13:20:40	U	SMPL
M1894-04LX5 - 1		4	0.0075	-	7470A	PPB	05 Apr 2021 13:22:55	U	SMPL
M1894-04A - 1		15930	4.1491	-	7470A	PPB	05 Apr 2021 13:25:11	U	SMPL
M1923-03LX5 - 1		-146	-0.0315	-	7470A	PPB	05 Apr 2021 13:27:26	U	SMPL
M1923-03A - 1		15069	3.9252	-	7470A	PPB	05 Apr 2021 13:29:42	U	SMPL
CCV38 - 1	CCV38	20201	5.2598	-	7470A	PPB	05 Apr 2021 13:31:57	U	SMPL
CCB38 - 1	CCB38	-99	-0.0193	-	7470A	PPB	05 Apr 2021 13:34:13	U	SMPL