



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113820

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/05/2021	13:15
2	CAL2	1	Water	NA	NA	20.2	7.00	04/05/2021	13:16
3	CAL3	1	Water	NA	NA	20.2	10.03	04/05/2021	13:18
4	ICV	1	Water	NA	NA	20.4	6.99	04/05/2021	13:20
5	CCV1	1	Water	NA	NA	20.3	2.01	04/05/2021	13:21
6	M1937-01	1	Water	NA	NA	20.4	5.21	04/05/2021	13:22
7	M1937-01DUP	1	Water	NA	NA	20.5	5.22	04/05/2021	13:23
8	CCV2	1	Water	NA	NA	20.3	12.02	04/05/2021	13:25

WORKLIST(Hardcopy Internal Chain)

LB 113820

WorkList Name : PH M1937 WorkList ID : 147734 Department : Wet-Chemistry Date : 04-05-2021 13:08:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/02/2021	Water	M1937-01	pH	Cool 4 deg C	STER16	L43	MAIN-CELL	04/02/2021	9040C

Date/Time 04/05/21 13:10
 Raw Sample Received by: JD (wet chem)
 Raw Sample Relinquished by: AM (wet chem)

Date/Time 04/05/21 16:20
 Raw Sample Received by: AM (wet chem)
 Raw Sample Relinquished by: JD (wet chem)