



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 4/6/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:30
In Date: 04/05/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 04/06/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113821

| Lab ID | Client SampleID | Dish # | Dish Wt (g) (A) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|-----------------|--------|-----------------|--------------------------|----------------------------|---------|----------|
| M1938-01 | MGB345 | 1 | 1.2 | 9.98 | 9.71 | 96.9 | |
| M1938-02 | MGB346 | 2 | 1.15 | 9.76 | 9.48 | 96.7 | |
| M1938-03 | MGB347 | 3 | 1.16 | 9.86 | 9.8 | 99.3 | |
| M1938-04 | MGB348 | 4 | 1.12 | 9.82 | 9.42 | 95.4 | |
| M1938-05 | MGB349 | 5 | 1.16 | 9.56 | 9.36 | 97.6 | |
| M1938-06 | MGB350 | 6 | 1.17 | 9.89 | 9.33 | 93.6 | |
| M1938-07 | MGB351 | 7 | 1.15 | 9.91 | 9.63 | 96.8 | |
| M1938-08 | MGB352 | 8 | 1.16 | 9.58 | 9.33 | 97.0 | |
| M1938-09 | MGB353 | 9 | 1.12 | 9.83 | 9.56 | 96.9 | |
| M1938-10 | MGB354 | 10 | 1.11 | 9.51 | 9.46 | 99.4 | |
| M1938-11 | MGB354D | 11 | 1.11 | 9.51 | 9.46 | 99.4 | |
| M1938-12 | MGB354S | 12 | 1.11 | 9.51 | 9.46 | 99.4 | |
| M1938-13 | MGB355 | 13 | 1.18 | 9.85 | 9.56 | 96.7 | |
| M1938-14 | MGB356 | 14 | 1.18 | 9.55 | 8.59 | 88.5 | |
| M1938-15 | MGB357 | 15 | 1.1 | 9.56 | 9.3 | 96.9 | |
| M1938-16 | MGB358 | 16 | 1.18 | 9.68 | 9.63 | 99.4 | |
| M1938-17 | MGB359 | 17 | 1.16 | 9.75 | 9.6 | 98.3 | |
| M1938-18 | MGB360 | 18 | 1.12 | 9.62 | 9.38 | 97.2 | |
| M1938-19 | MGB361 | 19 | 1.18 | 9.3 | 9.26 | 99.5 | |
| M1938-20 | MGB362 | 20 | 1.1 | 9.77 | 9.51 | 97.0 | |
| M1938-21 | MGB363 | 21 | 1.18 | 9.61 | 9.31 | 96.4 | |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$