

LB113824


7471B

INSTRUMENT ID: CV 1

Linear

μ Abs.:

58891



A= 0.0000e+000

B= 1.7092e-004

C= 1.6567e-002

Rho= 0.9997164

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.024	0.024	41	0.000	41					—
0.2	0.200	0.210	0.010	1130	0.0 %	1130					5
2.5	2.500	2.452	-0.048	14250	0.0 %	14250					-2
5.0	5.000	5.097	0.097	29726	0.0 %	29726					2
7.5	7.500	7.335	-0.165	42819	0.0 %	42819					-2
10.0	10.000	10.082	0.082	58891	0.0 %	58891					1

LB113824 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Apr 2021 14:22:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7471B	PPB	05 Apr 2021 14:25:06	S	Std
0.2 - 1	50.2	1130	-	0.2000	7471B	PPB	05 Apr 2021 14:27:22	S	Std
2.5 - 1	52.5	14250	-	2.5000	7471B	PPB	05 Apr 2021 14:29:39	S	Std
5.0 - 1	55	29726	-	5.0000	7471B	PPB	05 Apr 2021 14:31:56	S	Std
7.5 - 1	57.5	42819	-	7.5000	7471B	PPB	05 Apr 2021 14:34:12	S	Std
10.0 - 1	51v	58891	-	10.0000	7471B	PPB	05 Apr 2021 14:41:40	S	Std
ICV70 - 1	ICV70	23285	3.9964	-	7471B	PPB	05 Apr 2021 14:44:24	U	SMPL
ICB70 - 1	ICB70	-148	-0.0087	-	7471B	PPB	05 Apr 2021 14:46:39	U	SMPL
CCV39 - 1	CCV39	30189	5.1764	-	7471B	PPB	05 Apr 2021 14:48:55	U	SMPL
CCB39 - 1	CCB39	-165	-0.0116	-	7471B	PPB	05 Apr 2021 14:51:09	U	SMPL
CRA - 1	CRA	1091	0.2030	-	7471B	PPB	05 Apr 2021 14:53:26	U	SMPL
HighStd - 1	HighStd	58444	10.0057	-	7471B	PPB	05 Apr 2021 14:55:41	U	SMPL
ChkStd - 1	ChkStd	40551	6.9475	-	7471B	PPB	05 Apr 2021 14:57:56	U	SMPL
PB135548BL - 1	PBS	-182	-0.0145	-	7471B	PPB	05 Apr 2021 15:00:13	U	SMPL
PB135548BS - 1	LCSS	21562	3.7019	-	7471B	PPB	05 Apr 2021 15:05:05	U	SMPL
M1924-01 - 1	NB-07-040221	1158	0.2145	-	7471B	PPB	05 Apr 2021 15:12:16	U	SMPL
M1924-01DUP - 1	NB-07-040221DUP	1098	0.2042	-	7471B	PPB	05 Apr 2021 15:19:52	U	SMPL
M1924-01MS - 1	NB-07-040221MS	23313	4.0012	-	7471B	PPB	05 Apr 2021 15:24:35	U	SMPL
M1924-01MSD - 1	NB-07-040221MSD	23286	3.9966	-	7471B	PPB	05 Apr 2021 15:26:52	U	SMPL
M1932-01 - 1	TS-1	3471	0.6098	-	7471B	PPB	05 Apr 2021 15:29:07	U	SMPL
CCV40 - 1	CCV40	30637	5.2530	-	7471B	PPB	05 Apr 2021 15:31:22	U	SMPL
CCB40 - 1	CCB40	-110	-0.0022	-	7471B	PPB	05 Apr 2021 15:33:38	U	SMPL
M1932-02 - 1	TS-2	4388	0.7666	-	7471B	PPB	05 Apr 2021 15:35:53	U	SMPL
M1932-03 - 1	TS-3	3631	0.6372	-	7471B	PPB	05 Apr 2021 15:38:08	U	SMPL
M1924-01LX5 - 1		151	0.0424	-	7471B	PPB	05 Apr 2021 15:40:23	U	SMPL
M1924-01A - 1		24522	4.2078	-	7471B	PPB	05 Apr 2021 15:42:38	U	SMPL
CCV41 - 1	CCV41	30681	5.2605	-	7471B	PPB	05 Apr 2021 15:44:54	U	SMPL
CCB41 - 1	CCB41	-90	0.0012	-	7471B	PPB	05 Apr 2021 15:47:10	U	SMPL