



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113826

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	04/05/2021	09:45
2	M1903-01		1	10	100.00	>212.0	>212.0	04/05/2021	10:10
3	M1903-01DUP		1	10	100.00	>212.0	>212.0	04/05/2021	10:40
4	M1903-02		1	15	100.00	>212.0	>212.0	04/05/2021	11:05
5	M1903-03		1	12	100.00	>212.0	>212.0	04/05/2021	11:35
6	M1903-04		1	13	100.00	>212.0	>212.0	04/05/2021	11:58
7	M1903-05		1	14	100.00	>212.0	>212.0	04/05/2021	12:25
8	M1903-06		1	15	100.00	>212.0	>212.0	04/05/2021	12:45
9	M1903-07		1	14	100.00	>212.0	>212.0	04/05/2021	13:15
10	M1903-08		1	12	100.00	>212.0	>212.0	04/05/2021	13:40
11	M1903-09		1	11	100.00	>212.0	>212.0	04/05/2021	14:05
12	M1919-05		1	15	100.00	>212.0	>212.0	04/05/2021	14:32

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP4521

WorkList ID : 147728

Department : Wet-Chemistry

Date : 04-05-2021 09:17:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/08/2021	Water	M1903-01	Flash Point	Cool 4 deg C	PSEG03	O12	30776	04/01/2021	1010A
04/08/2021	Water	M1903-02	Flash Point	Cool 4 deg C	PSEG03	O12	30791	04/01/2021	1010A
04/08/2021	Water	M1903-03	Flash Point	Cool 4 deg C	PSEG03	O12	30792	04/01/2021	1010A
04/08/2021	Water	M1903-04	Flash Point	Cool 4 deg C	PSEG03	O12	41726	04/01/2021	1010A
04/08/2021	Water	M1903-05	Flash Point	Cool 4 deg C	PSEG03	O12	30809	04/01/2021	1010A
04/08/2021	Water	M1903-06	Flash Point	Cool 4 deg C	PSEG03	O12	41749	04/01/2021	1010A
04/08/2021	Water	M1903-07	Flash Point	Cool 4 deg C	PSEG03	O12	31029	04/01/2021	1010A
04/08/2021	Water	M1903-08	Flash Point	Cool 4 deg C	PSEG03	O12	31030	04/01/2021	1010A
04/08/2021	Water	M1903-09	Flash Point	Cool 4 deg C	PSEG03	O12	31031	04/01/2021	1010A
04/07/2021	Water	M1919-05	Flash Point	Cool 4 deg C	PSEG03	O13	23936	04/02/2021	1010A

Date/Time 04/05/21 09:25
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: RC Ext Lab

Date/Time 04/05/21 14:40
 Raw Sample Received by: RC Ext Lab
 Raw Sample Relinquished by: KS (Wetchem)