

LB113831


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

47001



A= 0.0000e+000

B= 2.1449e-004 *slope*

C= -4.1265e-003 *y-intercept*

Rho= 0.9998555

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>q/v1)</i>
0.00	0.000	0.012	0.012	74	0.000	74					-
0.05	0.050					204					-
0.20	0.200	0.197	-0.003	936	0.0 %	936					-2
2.50	2.500	2.518	0.018	11759	0.0 %	11759					1
5.00	5.000	5.025	0.025	23447	0.0 %	23447					1
7.50	7.500	7.371	-0.129	34386	0.0 %	34386					-2
10.0	10.000	10.077	0.077	47001	0.0 %	47001					1

LB113831

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	74	-		0 SFAM01.1	PPB	4/5/2021 16:35	Std
	0.2 S01	936	-		0.2 SFAM01.1	PPB	4/5/2021 16:37	Std
	2.5 S02	11759	-		2.5 SFAM01.1	PPB	4/5/2021 16:40	Std
	5 S03	23447	-		5 SFAM01.1	PPB	4/5/2021 16:42	Std
	7.5 S04	34386	-		7.5 SFAM01.1	PPB	4/5/2021 16:44	Std
	10 S05	47001	-		10 SFAM01.1	PPB	4/5/2021 16:46	Std
ICV018	ICV018	17804	3.8147 -		SFAM01.1	PPB	4/5/2021 16:49	SMPL
ICB018	ICB018	-70	-0.0191 -		SFAM01.1	PPB	4/5/2021 16:52	SMPL
CCV042	CCV42	23569	5.0512 -		SFAM01.1	PPB	4/5/2021 16:54	SMPL
CCB042	CCB42	-93	-0.0241 -		SFAM01.1	PPB	4/5/2021 16:56	SMPL
PB135549BL	PBW549	2	-0.0037 -		SFAM01.1	PPB	4/5/2021 16:58	SMPL
M1904-07	X3246	13226	2.8327 -		SFAM01.1	PPB	4/5/2021 17:01	SMPL
M1904-08	X3246D	13238	2.8353 -		SFAM01.1	PPB	4/5/2021 17:07	SMPL
M1904-09	X3246S	17896	3.8344 -		SFAM01.1	PPB	4/5/2021 17:10	SMPL
CCV043	CCV043	23279	4.989 -		SFAM01.1	PPB	4/5/2021 17:14	SMPL
CCB043	CCB043	-120	-0.0299 -		SFAM01.1	PPB	4/5/2021 17:17	SMPL