



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113832

Slope : 97.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/05/2021	16:45
2	CAL2	1	Water	NA	NA	20.2	7.00	04/05/2021	16:46
3	CAL3	1	Water	NA	NA	20.2	10.03	04/05/2021	16:47
4	ICV	1	Water	NA	NA	20.2	6.99	04/05/2021	16:48
5	CCV1	1	Water	NA	NA	20.3	2.01	04/05/2021	16:49
6	M1934-01	1	Solid	20.02	20	21.7	6.36	04/05/2021	17:00
7	M1934-01DUP	1	Solid	20.03	20	21.8	6.37	04/05/2021	17:01
8	M1934-03	1	Solid	20.03	20	21.6	7.76	04/05/2021	17:03
9	M1934-05	1	Solid	20.04	20	21.7	8.76	04/05/2021	17:05
10	CCV2	1	Water	NA	NA	20.6	12.02	04/05/2021	17:07

WORKLIST(Hardcopy Internal Chain)

VB 113832

WorkList Name : CORROSIONITY M1934 WorkList ID : 147746 Department : Wet-Chemistry Date : 04-05-2021 15:42:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/05/2021	Solid	M1934-01	Corrosivity	Cool 4 deg C	PSEG03	O22	TP-3	04/05/2021	9045D
04/05/2021	Solid	M1934-03	Corrosivity	Cool 4 deg C	PSEG03	O22	TP-2	04/05/2021	9045D
04/05/2021	Solid	M1934-05	Corrosivity	Cool 4 deg C	PSEG03	O22	TP-1	04/05/2021	9045D

Date/Time 04/05/21 16:10
 Raw Sample Received by: JP (wet chem)
 Raw Sample Relinquished by: KS (wet chem)

Date/Time 04/05/21 17:15
 Raw Sample Received by: KS (wet chem)
 Raw Sample Relinquished by: JP (wet chem)