

LB113838


7471B

INSTRUMENT ID: CVL

Linear

μ Abs.:

53705



A= 0.0000e+000

B= 1.8339e-004

C= 2.0787e-002

Rho= 0.9995789

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.023	0.023	12	0.000	12					-
0.2	0.200	0.193	-0.007	940	0.0 %	940					-4
2.5	2.500	2.360	-0.140	12758	0.0 %	12758					-6
5.0	5.000	5.128	0.128	27849	0.0 %	27849					3
7.5	7.500	7.625	0.125	41464	0.0 %	41464					2
10.0	10.000	9.870	-0.130	53709	0.0 %	53709					-1

LB113838 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Apr 2021 12:20:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	12	-	0.0000	7471B	PPB	06 Apr 2021 12:22:05	S	Std
0.2 - 1	50.2	940	-	0.2000	7471B	PPB	06 Apr 2021 12:24:24	S	Std
2.5 - 1	52.5	12758	-	2.5000	7471B	PPB	06 Apr 2021 12:26:42	S	Std
5.0 - 1	55	27849	-	5.0000	7471B	PPB	06 Apr 2021 12:28:58	S	Std
7.5 - 1	57.5	41464	-	7.5000	7471B	PPB	06 Apr 2021 12:31:14	S	Std
10.0 - 1	59	53709	-	10.0000	7471B	PPB	06 Apr 2021 12:33:31	S	Std
ICV72 - 1	ICV72	21482	3.9604	-	7471B	PPB	06 Apr 2021 12:38:59	U	SMPL
ICB72 - 1	ICB72	-192	-0.0144	-	7471B	PPB	06 Apr 2021 12:41:14	U	SMPL
CCV44 - 1	CCV44	28609	5.2674	-	7471B	PPB	06 Apr 2021 12:43:30	U	SMPL
CCB44 - 1	CCB44	-211	-0.0179	-	7471B	PPB	06 Apr 2021 12:45:44	U	SMPL
CRA - 1	CRA	918	0.1891	-	7471B	PPB	06 Apr 2021 12:47:59	U	SMPL
HighStd - 1	HighStd	54520	10.0192	-	7471B	PPB	06 Apr 2021 12:50:14	U	SMPL
ChkStd - 1	ChkStd	39943	7.3459	-	7471B	PPB	06 Apr 2021 12:52:30	U	SMPL
PB135565BL - 1	PBS	-242	-0.0236	-	7471B	PPB	06 Apr 2021 12:54:48	U	SMPL
PB135565BS - 1	LCSS	20303	3.7442	-	7471B	PPB	06 Apr 2021 12:57:05	U	SMPL
M1933-02 - 1	BENNETT_BH_01_0-1	15904	2.9374	-	7471B	PPB	06 Apr 2021 13:04:10	U	SMPL
M1933-02DUP - 1	BENNETT_BH_01_0-1DUP	15761	2.9112	-	7471B	PPB	06 Apr 2021 13:06:26	U	SMPL
M1933-02MS - 1	BENNETT_BH_01_0-1MS	34265	6.3046	-	7471B	PPB	06 Apr 2021 13:08:43	U	SMPL
M1933-02MSD - 1	BENNETT_BH_01_0-1MSD	36134	6.6474	-	7471B	PPB	06 Apr 2021 13:10:59	U	SMPL
M1933-04 - 1	BENNETT_BH_01_1-1.5	6478	1.2088	-	7471B	PPB	06 Apr 2021 13:13:14	U	SMPL
CCV45 - 1	CCV45	28692	5.2826	-	7471B	PPB	06 Apr 2021 13:15:30	U	SMPL
CCB45 - 1	CCB45	-20	0.0171	-	7471B	PPB	06 Apr 2021 13:17:46	U	SMPL
M1933-06 - 1	BENNETT_BH_01_2-2.5	3708	0.7008	-	7471B	PPB	06 Apr 2021 13:20:01	U	SMPL
M1933-08 - 1	BENNETT_BH_04_0-1	6142	1.1472	-	7471B	PPB	06 Apr 2021 13:22:16	U	SMPL
M1933-10 - 1	BENNETT_BH_04_1-1.5	14170	2.6194	-	7471B	PPB	06 Apr 2021 13:24:32	U	SMPL
M1933-12 - 1	BENNETT_BH_04_2-2.5	13089	2.4212	-	7471B	PPB	06 Apr 2021 13:26:47	U	SMPL
M1933-14 - 1	BENNETT_BH_05_0-1	5936	1.1094	-	7471B	PPB	06 Apr 2021 13:29:03	U	SMPL
M1933-16 - 1	BENNETT_BH_05_1-1.5	9894	1.8352	-	7471B	PPB	06 Apr 2021 13:31:19	U	SMPL
M1933-18 - 1	BENNETT_BH_05_2-2.5	4251	0.8004	-	7471B	PPB	06 Apr 2021 13:33:35	U	SMPL
M1934-01 - 1	TP-3	709	0.1508	-	7471B	PPB	06 Apr 2021 13:35:52	U	SMPL
M1934-03 - 1	TP-2	1121	0.2264	-	7471B	PPB	06 Apr 2021 13:38:08	U	SMPL
M1934-05 - 1	TP-1	679	0.1453	-	7471B	PPB	06 Apr 2021 13:40:24	U	SMPL
CCV46 - 1	CCV46	29310	5.3959	-	7471B	PPB	06 Apr 2021 13:42:39	U	SMPL
CCB46 - 1	CCB46	-1	0.0206	-	7471B	PPB	06 Apr 2021 13:44:55	U	SMPL
M1933-02LX5 - 1		2951	0.5620	-	7471B	PPB	06 Apr 2021 13:47:10	U	SMPL
M1933-02A - 1		38007	6.9909	-	7471B	PPB	06 Apr 2021 13:49:25	U	SMPL
CCV47 - 1	CCV47	29400	5.4125	-	7471B	PPB	06 Apr 2021 13:51:41	U	SMPL
CCB47 - 1	CCB47	-29	0.0155	-	7471B	PPB	06 Apr 2021 13:53:56	U	SMPL