


LB113845

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

5002E



Linear

A= 0.0000e+000

B= 2.0084e-004 slope

C= 3.0598e-002 intercept

Rho= 0.9998646

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/c D
0.00	0.000	0.044	0.044	65	0.000	65					
0.05	0.050					204					
0.20	0.200	0.188	-0.012	784	0.0 %	784					-6
2.50	2.500	2.484	-0.016	12215	0.0 %	12215					-1
5.00	5.000	5.023	0.023	24856	0.0 %	24856					0
7.50	7.500	7.385	-0.115	36617	0.0 %	36617					-2
10.0	10.000	10.077	0.077	50025	0.0 %	50025					1

LB113845

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	65	-		0 SFAM01.1	PPB	4/6/2021 14:58	Std
	0.2 S01	784	-		0.2 SFAM01.1	PPB	4/6/2021 15:00	Std
	2.5 S02	12215	-		2.5 SFAM01.1	PPB	4/6/2021 15:02	Std
	5 S03	24856	-		5 SFAM01.1	PPB	4/6/2021 15:05	Std
	7.5 S04	36617	-		7.5 SFAM01.1	PPB	4/6/2021 15:07	Std
	10 S05	50025	-		10 SFAM01.1	PPB	4/6/2021 15:09	Std
ICV019	ICV019	18436	3.7332 -		SFAM01.1	PPB	4/6/2021 15:14	SMPL
ICB019	ICB019	-111	0.0083 -		SFAM01.1	PPB	4/6/2021 15:17	SMPL
CCV048	CCV048	25170	5.0856 -		SFAM01.1	PPB	4/6/2021 15:19	SMPL
CCB048	CCB048	-124	0.0057 -		SFAM01.1	PPB	4/6/2021 15:21	SMPL
PB135566BL	PBS566	15	0.0336 -		SFAM01.1	PPB	4/6/2021 15:23	SMPL
M1904-19	X3250	10018	2.0426 -		SFAM01.1	PPB	4/6/2021 15:28	SMPL
M1904-20	X3250D	10946	2.2289 -		SFAM01.1	PPB	4/6/2021 15:32	SMPL
M1904-21	X3250S	21971	4.4431 -		SFAM01.1	PPB	4/6/2021 15:35	SMPL
CCV049	CCV049	24936	5.0386 -		SFAM01.1	PPB	4/6/2021 15:39	SMPL
CCB049	CCB049	-81	0.0143 -		SFAM01.1	PPB	4/6/2021 15:41	SMPL