



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 4/7/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:35
In Date: 04/06/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:28
Out Date: 04/07/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113847

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M1886-11	RO-124037	1	1.13	9.69	7.8	77.9	
M1952-01	17210	14	1.14	9.78	8.22	81.9	
M1953-01	HD-02-040621	15	1.18	9.65	8.92	91.4	
M1953-02	HD-02-040621-EPH-2	16	1.16	9.78	9.06	91.6	
M1954-01	SU-03-040621	17	1.16	9.79	9.16	92.7	
M1954-02	SU-03-040621-EPH-2	18	1.15	9.51	8.85	92.1	
M1954-02D	SU-03-040621-EPH-2DUP	19	1.15	9.53	8.87	92.1	
M1955-01	SIDEWALK-206	20	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M1955-02	SIDEWALK-206-EPH-2	21	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M1956-01	OIL-1-2-3-4-5	22	1.00	2.00	2.00	100.0	OIL sample
M1956-02	OIL-DEBRIS-3-5	23	1.00	2.00	2.00	100.0	OILY-DEBRIS
M1961-01	TR1-A-D-COMP	2	1.15	9.86	7.67	74.9	
M1961-03	TR1-A-BOTTOM	3	1.16	9.6	6.8	66.8	
M1961-12	TR2-A-D-COMP	4	1.19	9.62	6.91	67.9	
M1961-14	TR2-A-BOTTOM	5	1.19	9.66	7.9	79.2	
M1963-01	S-1	6	1.14	9.97	9.16	90.8	
M1963-02	S-1A	7	1.18	9.74	8.91	90.3	
M1963-03	S-2	8	1.18	9.97	8.6	84.4	
M1963-04	S-2A	9	1.14	9.56	7.91	80.4	
M1963-05	S-3	10	1.16	9.8	7.45	72.8	
M1963-06	S-3A	11	1.16	9.89	7.58	73.5	
M1963-07	S-4	12	1.15	9.71	8.94	91.0	
M1963-08	S-4A	13	1.19	9.85	8.98	90.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$