



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113851

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/06/2021	16:30
2	CAL2	1	Water	NA	NA	20.2	7.00	04/06/2021	16:31
3	CAL3	1	Water	NA	NA	20.2	10.03	04/06/2021	16:32
4	ICV	1	Water	NA	NA	20.3	6.99	04/06/2021	16:35
5	CCV1	1	Water	NA	NA	20.3	2.01	04/06/2021	16:36
6	M1952-02	1	Water	NA	NA	20.2	6.38	04/06/2021	16:38
7	M1952-03	1	Water	NA	NA	20.6	5.39	04/06/2021	16:40
8	M1952-04	1	Water	NA	NA	20.6	6.85	04/06/2021	16:42
9	M1955-03	1	Water	NA	NA	20.4	6.28	04/06/2021	16:45
10	M1955-03DUP	1	Water	NA	NA	20.5	6.29	04/06/2021	16:46
11	CCV2	1	Water	NA	NA	20.3	12.02	04/06/2021	16:48

WORKLIST(Hardcopy Internal Chain)

18 113851

WorkList Name : ph m1953 WorkList ID : 147784 Department : Wet-Chemistry Date : 04-06-2021 16:14:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/06/2021	Water	M1952-02	pH	Cool 4 deg C	PSEG03	O23	41738	04/06/2021	9040C
04/06/2021	Water	M1952-03	pH	Cool 4 deg C	PSEG03	O23	30820	04/06/2021	9040C
04/06/2021	Water	M1952-04	pH	Cool 4 deg C	PSEG03	O23	31027	04/06/2021	9040C
04/06/2021	Water	M1955-03	pH	Cool 4 deg C	PSEG03	O23	2885	04/06/2021	9040C

Date/Time 04/06/21 16:20
 Raw Sample Received by: JO (Wet-Chem)
 Raw Sample Relinquished by: KS (Wet-Chem)

Date/Time 04/06/21 17:30
 Raw Sample Received by: KS (Wet-Chem)
 Raw Sample Relinquished by: JO (Wet-Chem)