

LB 113 854


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37416



A= 0.0000e+000

B= 2.6531e-004

C= -1.6465e-002

Rho= 0.9998419

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.017	-0.017	-2	0.000	-2					-
0.2	0.200	0.177	-0.023	729	0.0 %	729					-12
2.5	2.500	2.463	-0.037	9346	0.0 %	9346					-1
5.0	5.000	5.113	0.113	19333	0.0 %	19333					2
7.5	7.500	7.554	0.054	28534	0.0 %	28534					1
10.0	10.000	9.910	-0.090	37416	0.0 %	37416					-1

LB113854 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 07 Apr 2021 09:37:06					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-2	-	0.0000	7470A	PPB	07 Apr 2021 09:39:49	S	Std
0.2 - 1	50.2	729	-	0.2000	7470A	PPB	07 Apr 2021 09:42:06	S	Std
2.5 - 1	52.5	9346	-	2.5000	7470A	PPB	07 Apr 2021 09:44:23	S	Std
5.0 - 1	55	19333	-	5.0000	7470A	PPB	07 Apr 2021 09:46:40	S	Std
7.5 - 1	57.5	28534	-	7.5000	7470A	PPB	07 Apr 2021 09:48:56	S	Std
10.0 - 1	510	37416	-	10.0000	7470A	PPB	07 Apr 2021 09:51:13	S	Std
ICV74 - 1	ICV74	16067	4.2463	-	7470A	PPB	07 Apr 2021 09:56:35	U	SMPL
ICB74 - 1	ICB74	-55	-0.0311	-	7470A	PPB	07 Apr 2021 10:01:26	U	SMPL
CCV54 - 1	CCV54	19317	5.1085	-	7470A	PPB	07 Apr 2021 10:03:41	U	SMPL
CCB54 - 1	CCB54	55	-0.0019	-	7470A	PPB	07 Apr 2021 10:08:43	U	SMPL
CRA - 1	CRA	712	0.1724	-	7470A	PPB	07 Apr 2021 10:10:58	U	SMPL
HighStd - 1	HighStd	38031	10.0735	-	7470A	PPB	07 Apr 2021 10:14:46	U	SMPL
ChkStd - 1	ChkStd	25374	6.7155	-	7470A	PPB	07 Apr 2021 10:17:01	U	SMPL
PB135573BL - 1	PBW	151	0.0236	-	7470A	PPB	07 Apr 2021 10:19:18	U	SMPL
PB135573BS - 1	LCSW	16058	4.2439	-	7470A	PPB	07 Apr 2021 10:24:09	U	SMPL
M1934-02 - 1	TP-3	202	0.0371	-	7470A	PPB	07 Apr 2021 10:26:24	U	SMPL
M1934-02DUP - 1	TP-3DUP	343	0.0745	-	7470A	PPB	07 Apr 2021 10:28:40	U	SMPL
M1934-02MS - 1	TP-3MS	16145	4.2669	-	7470A	PPB	07 Apr 2021 10:33:53	U	SMPL
M1934-02MSD - 1	TP-3MSD	16368	4.3261	-	7470A	PPB	07 Apr 2021 10:36:10	U	SMPL
M1934-04 - 1	TP-2	93	0.0082	-	7470A	PPB	07 Apr 2021 10:38:24	U	SMPL
CCV55 - 1	CCV55	20023	5.2958	-	7470A	PPB	07 Apr 2021 10:40:40	U	SMPL
CCB55 - 1	CCB55	-256	-0.0844	-	7470A	PPB	07 Apr 2021 10:42:55	U	SMPL
M1934-06 - 1	TP-1	312	0.0663	-	7470A	PPB	07 Apr 2021 10:45:11	U	SMPL
M1935-01 - 1	336	243	0.0480	-	7470A	PPB	07 Apr 2021 10:47:27	U	SMPL
PB135574BL - 1	PBW	174	0.0297	-	7470A	PPB	07 Apr 2021 10:49:42	U	SMPL
PB135574BS - 1	LCSW	14334	3.7865	-	7470A	PPB	07 Apr 2021 10:51:58	U	SMPL
M1919-05 - 1	23936	1743	0.4460	-	7470A	PPB	07 Apr 2021 10:54:15	U	SMPL
M1949-01 - 1	AR-1	362	0.0796	-	7470A	PPB	07 Apr 2021 10:56:33	U	SMPL
M1949-02 - 1	PE-1	429	0.0974	-	7470A	PPB	07 Apr 2021 10:58:50	U	SMPL
M1949-03 - 1	MA-1	447	0.1021	-	7470A	PPB	07 Apr 2021 11:01:07	U	SMPL
M1949-04 - 1	CA-1	451	0.1032	-	7470A	PPB	07 Apr 2021 11:03:23	U	SMPL
M1949-05 - 1	AB-1	442	0.1008	-	7470A	PPB	07 Apr 2021 11:05:38	U	SMPL
CCV56 - 1	CCV56	19118	5.0557	-	7470A	PPB	07 Apr 2021 11:07:54	U	SMPL
CCB56 - 1	CCB56	-63	-0.0332	-	7470A	PPB	07 Apr 2021 11:10:09	U	SMPL
M1951-01 - 1	SW-1	473	0.1090	-	7470A	PPB	07 Apr 2021 11:12:25	U	SMPL
M1951-01DUP - 1	SW-1DUP	454	0.1040	-	7470A	PPB	07 Apr 2021 11:14:40	U	SMPL
M1951-02 - 1	M1951-01MS	15377	4.0632	-	7470A	PPB	07 Apr 2021 11:16:56	U	SMPL
M1951-03 - 1	M1951-01MSD	15553	4.1099	-	7470A	PPB	07 Apr 2021 11:19:11	U	SMPL
M1951-04 - 1	SW-2	296	0.0621	-	7470A	PPB	07 Apr 2021 11:21:27	U	SMPL
M1951-05 - 1	EB-2021-04-05_SW	81	0.0050	-	7470A	PPB	07 Apr 2021 11:23:44	U	SMPL
CCV57 - 1	CCV57	18746	4.9570	-	7470A	PPB	07 Apr 2021 11:26:00	U	SMPL
CCB57 - 1	CCB57	-176	-0.0632	-	7470A	PPB	07 Apr 2021 11:28:17	U	SMPL
PB135573TB - 1	PB135550TB	150	0.0233	-	7470A	PPB	07 Apr 2021 11:33:09	U	SMPL
M1934-02LX5 - 1		61	-0.0003	-	7470A	PPB	07 Apr 2021 11:35:24	U	SMPL
M1934-02A - 1		14626	3.8639	-	7470A	PPB	07 Apr 2021 11:37:40	U	SMPL
M1951-01LX5 - 1		-93	-0.0411	-	7470A	PPB	07 Apr 2021 11:39:55	U	SMPL
M1951-01A - 1		16199	4.2813	-	7470A	PPB	07 Apr 2021 11:42:11	U	SMPL
CCV58 - 1	CCV58	19508	5.1592	-	7470A	PPB	07 Apr 2021 11:44:26	U	SMPL
CCB58 - 1	CCB58	-165	-0.0602	-	7470A	PPB	07 Apr 2021 11:46:42	U	SMPL