



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB113859

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP90379
Chlorine Calibration std, 0.1ppm	WP90374
Chlorine Calibration std, 0.2ppm	WP90375
Chlorine Calibration std, 0.8ppm	WP90377
Chlorine Calibration std, 0.0ppm	WP90373
Chlorine Calibration std, 1.6ppm	WP90378
Residual Chlorine Calibration and CCV std, 0	WP90376
Total Chlorine Powder Pillows	W2249

Intercept: 0.0111

Slope: 1.0591

Regression: 0.999908

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		04/07/2021	10:48
2	CAL2	0.1	1	0.000	0.120	0.120	0.10	-3	04/07/2021	10:50
3	CAL3	0.2	1	0.000	0.230	0.230	0.21	-3.5	04/07/2021	10:51
4	CAL4	0.4	1	0.000	0.430	0.430	0.40	1	04/07/2021	10:53
5	CAL5	0.8	1	0.000	0.870	0.870	0.81	-1.4	04/07/2021	10:55
6	CAL6	1.6	1	0.000	1.700	1.700	1.60	0.3	04/07/2021	10:57



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB113859

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4400	0.4400	0.3690	04/07/2021	10:59
2	ICB				1	0.0000	0.0100	0.0100	-0.0010	04/07/2021	11:01
3	CCV1			0.4	1	0.0000	0.4400	0.4400	0.3690	04/07/2021	11:02
4	CCB1				1	0.0000	0.0100	0.0100	-0.0510	04/07/2021	11:04
5	LB113859BL	50	50		1	0.0000	0.0100	0.0100	-0.0510	04/07/2021	11:06
6	LB113859BS	50	50	0.4	1	0.0000	0.4500	0.4500	0.3780	04/07/2021	11:08
7	M1965-01	50	50		1	0.0000	0.0100	0.0100	-0.0510	04/07/2021	11:10
8	M1965-01DUP	50	50		1	0.0000	0.0100	0.0100	-0.0510	04/07/2021	11:12
9	CCV2			0.4	1	0.0000	0.4500	0.4500	0.3780	04/07/2021	11:15
10	CCB2				1	0.0000	0.0000	0.0000	-0.0610	04/07/2021	11:17

LB113859
WORKLIST(Hardcopy Internal Chain)

WorkList Name : ResidualChlorine-40721 WorkList ID : 147805 Department : Wet-Chemistry Date : 04-07-2021 09:14:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/21/2021	Water	M1965-01 A	Residual Chlorine	Cool 4 deg C	STER17	O13	JS9600	03/21/2021	SM4500 Cl G

Date/Time 04/07/2021 10:20
Raw Sample Received by: AM(wet-chem)
Raw Sample Relinquished by: SP(wet-chem)

Date/Time 04/07/2021 11:30
Raw Sample Received by: SP(wet-chem)
Raw Sample Relinquished by: AM(wet-chem)