



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 4/8/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/07/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:28
Out Date: 04/08/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113861

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M1765-07	RBR-54	4	1.16	9.85	7.66	74.8	
M1933-01	BENNETT_BH_01_0-1	1	1.15	9.96	8.55	84.0	
M1969-01	OK-01-040721	5	1.15	9.96	9.47	94.4	
M1969-02	OK-01-040721-EPH-2	6	1.15	9.55	9.00	93.5	
M1970-01	JC-03-040721	7	1.16	9.64	9.05	93.0	
M1970-02	JC-03-040721-EPH-2	8	1.17	9.69	9.11	93.2	
M1971-01	HR-02-040721	9	1.14	9.55	9.01	93.6	
M1971-02	HR-02-040721-EPH-2	10	1.17	9.75	9.16	93.1	
M1971-03	HR-03-040721	11	1.16	9.7	9.53	98.0	
M1971-04	HR-03-040721-EPH-2	12	1.17	9.95	9.6	96.0	
M1972-01	ETGI-371	2	1.15	9.6	8.73	89.7	
M1972-02	ETGI-371	3	1.19	9.93	8.96	88.9	
M1973-01	BU-02-040721	13	1.18	9.61	9.14	94.4	
M1975-01	RBR-200032	14	1.18	9.55	8.86	91.8	
M1975-02	RBR-200032-EPH-2	15	1.16	9.66	8.24	83.3	
M1979-01	S-1	16	1.15	9.78	8.35	83.4	
M1979-02	S-2	17	1.13	9.96	8.82	87.1	
M1979-03	S-3	18	1.16	9.64	8.14	82.3	
M1979-04	S-4	19	1.15	9.98	8.85	87.2	
M1979-05	S-5	20	1.16	9.87	8.59	85.3	
M1979-06	S-6	21	1.14	9.79	8.44	84.4	
M1979-07	S-7	22	1.14	9.96	8.65	85.1	
M1979-08	S-8	23	1.14	9.95	8.97	88.9	
M1979-09	S-9	24	1.15	9.96	9.00	89.1	
M1979-10	S-10	25	1.15	9.57	8.16	83.3	
M1979-10D	S-10DUP	26	1.16	9.59	8.18	83.3	
M1979-11	S-11	27	1.15	9.65	9.04	92.8	
M1979-12	S-12	28	1.15	9.54	8.94	92.8	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 4/8/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/07/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:28
Out Date: 04/08/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113861

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M1979-13	S-13	29	1.16	9.78	8.42	84.2	
M1979-14	S-14	30	1.14	9.97	8.68	85.4	
M1979-15	S-15	31	1.15	9.93	9.26	92.4	
M1979-16	S-16	32	1.16	9.67	8.26	83.4	
M1979-17	S-17	33	1.15	9.95	8.69	85.7	
M1979-18	S-18	34	1.14	9.98	8.64	84.8	
M1979-19	S-19	35	1.15	9.93	8.45	83.1	
M1979-20	S-20	36	1.16	9.68	8.09	81.3	
M1979-21	S-21	37	1.16	9.94	8.48	83.4	
M1979-22	S-22	38	1.15	9.76	8.95	90.6	
M1979-23	S-23	39	1.15	9.87	8.34	82.5	
M1980-01	B8	40	1.17	9.97	9.05	89.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$