


LB 113865

7471B

INSTRUMENT ID: CV1

 μ Abs.:

55374



Linear

A= 0.0000e+000

B= 1.8140e-004

C= -2.0355e-002

Rho= 0.9997917

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Val
0.0	0.000	0.019	0.019	217	0.000	217					
0.2	0.200	0.174	-0.026	1074	0.0 %	1074					
2.5	2.500	2.601	0.101	14451	0.0 %	14451					
5.0	5.000	4.854	-0.146	26868	0.0 %	26868					
7.5	7.500	7.527	0.027	41607	0.0 %	41607					
10.0	10.000	10.025	0.025	55374	0.0 %	55374					

Val
-13
4
-3
0
0

LB113865 INSTRUMENT ID :: CV1

Method: 7471B		Operator: Admin		Date of Analysis: 07 Apr 2021 15:03:42					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	217	-	0.0000	7471B	PPB	07 Apr 2021 15:06:17	S	Std
0.2 - 1	50.2	1074	-	0.2000	7471B	PPB	07 Apr 2021 15:08:34	S	Std
2.5 - 1	52.5	14451	-	2.5000	7471B	PPB	07 Apr 2021 15:10:51	S	Std
5.0 - 1	55	26868	-	5.0000	7471B	PPB	07 Apr 2021 15:13:08	S	Std
7.5 - 1	57.5	41607	-	7.5000	7471B	PPB	07 Apr 2021 15:15:24	S	Std
10.0 - 1	510	55374	-	10.0000	7471B	PPB	07 Apr 2021 15:20:17	S	Std
ICV75 - 1	ICV75	22830	4.1211	-	7471B	PPB	07 Apr 2021 15:25:27	U	SMPL
ICB75 - 1	ICB75	-6	-0.0214	-	7471B	PPB	07 Apr 2021 15:27:42	U	SMPL
CCV59 - 1	CCV59	27102	4.8960	-	7471B	PPB	07 Apr 2021 15:29:59	U	SMPL
CCB59 - 1	CCB59	-51	-0.0296	-	7471B	PPB	07 Apr 2021 15:32:14	U	SMPL
CRA - 1	CRA	1036	0.1676	-	7471B	PPB	07 Apr 2021 15:34:31	U	SMPL
HighStd - 1	HighStd	56241	10.1819	-	7471B	PPB	07 Apr 2021 15:36:45	U	SMPL
ChkStd - 1	ChkStd	38304	6.9281	-	7471B	PPB	07 Apr 2021 15:39:00	U	SMPL
PB135601BL - 1	PB135601BL	-110	-0.0403	-	7471B	PPB	07 Apr 2021 15:41:18	U	SMPL
PB135601BS - 1	PB135601BS	21088	3.8051	-	7471B	PPB	07 Apr 2021 15:46:34	U	SMPL
M1952-01 - 1	17210	2798	0.4872	-	7471B	PPB	07 Apr 2021 15:48:50	U	SMPL
M1953-01 - 1	HD-02-040621	4776	0.8460	-	7471B	PPB	07 Apr 2021 15:51:07	U	SMPL
M1954-01 - 1	SU-03-040621	2757	0.4798	-	7471B	PPB	07 Apr 2021 15:53:25	U	SMPL
M1955-01 - 1	SIDEWALK-206	2557	0.4435	-	7471B	PPB	07 Apr 2021 15:55:43	U	SMPL
M1956-02 - 1	OIL-DEBRIS-3-5	133	0.0038	-	7471B	PPB	07 Apr 2021 15:57:59	U	SMPL
CCV60 - 1	CCV60	26238	4.7393	-	7471B	PPB	07 Apr 2021 16:00:14	U	SMPL
CCB60 - 1	CCB60	96	-0.0029	-	7471B	PPB	07 Apr 2021 16:02:29	U	SMPL
M1961-01 - 1	TR1-A-D-COMP	2741	0.4769	-	7471B	PPB	07 Apr 2021 16:04:44	U	SMPL
M1961-12 - 1	TR2-A-D-COMP	4737	0.8389	-	7471B	PPB	07 Apr 2021 16:07:00	U	SMPL
M1963-01 - 1	S-1	1344	0.2234	-	7471B	PPB	07 Apr 2021 16:09:17	U	SMPL
M1963-03 - 1	S-2	4300	0.7597	-	7471B	PPB	07 Apr 2021 16:11:33	U	SMPL
M1963-05 - 1	S-3	8867	1.5881	-	7471B	PPB	07 Apr 2021 16:13:48	U	SMPL
M1963-07 - 1	S-4	839	0.1318	-	7471B	PPB	07 Apr 2021 16:16:04	U	SMPL
M1963-07DUP - 1	S-4DUP	695	0.1057	-	7471B	PPB	07 Apr 2021 16:20:54	U	SMPL
M1963-07MS - 1	S-4MS	22952	4.1432	-	7471B	PPB	07 Apr 2021 16:28:40	U	SMPL
M1963-07MSD - 1	S-4MSD	23033	4.1579	-	7471B	PPB	07 Apr 2021 16:30:57	U	SMPL
CCV61 - 1	CCV61	26684	4.8202	-	7471B	PPB	07 Apr 2021 16:33:13	U	SMPL
CCB61 - 1	CCB61	36	-0.0138	-	7471B	PPB	07 Apr 2021 16:35:28	U	SMPL
M1963-07LX5 - 1		648	0.0972	-	7471B	PPB	07 Apr 2021 16:37:43	U	SMPL
M1963-07A - 1		23405	4.2254	-	7471B	PPB	07 Apr 2021 16:39:59	U	SMPL
CCV62 - 1	CCV62	26881	4.8559	-	7471B	PPB	07 Apr 2021 16:42:14	U	SMPL
CCB62 - 1	CCB62	38	-0.0135	-	7471B	PPB	07 Apr 2021 16:44:29	U	SMPL