



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113867

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/07/2021	17:00
2	CAL2	1	Water	NA	NA	20.7	7.00	04/07/2021	17:01
3	CAL3	1	Water	NA	NA	20.6	10.03	04/07/2021	17:02
4	ICV	1	Water	NA	NA	20.6	6.99	04/07/2021	17:05
5	CCV1	1	Water	NA	NA	20.6	2.01	04/07/2021	17:06
6	M1973-01	1	Solid	20.03	20	20.9	5.07	04/07/2021	17:09
7	M1973-01DUP	1	Solid	20.04	20	20.8	5.08	04/07/2021	17:10
8	CCV2	1	Water	NA	NA	20.7	12.02	04/07/2021	17:11

WORKLIST(Hardcopy Internal Chain)

HB 113867

WorkList Name : PH M1973 S WorkList ID : 147832 Department : Wet-Chemistry Date : 04-07-2021 16:42:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/07/2021	Solid	M1973-01	pH	Cool 4 deg C	PSEG05	O32	BU-02-040721	04/07/2021	9045D

Date/Time 04/07/21 16:45
 Raw Sample Received by: JP Wet Chem
 Raw Sample Relinquished by: KS (Wet Chem)

Date/Time 04/07/21 17:30
 Raw Sample Received by: KS (Wet Chem)
 Raw Sample Relinquished by: JP Wet Chem