

LB113873


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37585



A= 0.0000e+000

B= 2.6453e-004

C= 1.0297e-002

Rho= 0.9999029

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.014	0.014	13	0.000	13					-
0.2	0.200	0.181	-0.019	645	0.0 %	645					-10
2.5	2.500	2.517	0.017	9475	0.0 %	9475					1
5.0	5.000	4.940	-0.060	18636	0.0 %	18636					-1
7.5	7.500	7.596	0.096	28676	0.0 %	28676					1
10.0	10.000	9.953	-0.047	37585	0.0 %	37585					0

LB113873 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Apr 2021 09:32:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	13	-	0.0000	7470A	PPB	08 Apr 2021 09:46:25	S	Std
0.2 - 1	50.2	645	-	0.2000	7470A	PPB	08 Apr 2021 09:48:42	S	Std
2.5 - 1	52.5	9475	-	2.5000	7470A	PPB	08 Apr 2021 09:50:58	S	Std
5.0 - 1	55	18636	-	5.0000	7470A	PPB	08 Apr 2021 09:53:15	S	Std
7.5 - 1	57.5	28676	-	7.5000	7470A	PPB	08 Apr 2021 09:58:05	S	Std
10.0 - 1	510	37585	-	10.0000	7470A	PPB	08 Apr 2021 10:05:29	S	Std
ICV76 - 1	ICV76	15557	4.1256	-	7470A	PPB	08 Apr 2021 10:10:37	U	SMPL
ICB76 - 1	ICB76	-127	-0.0233	-	7470A	PPB	08 Apr 2021 10:12:52	U	SMPL
CCV63 - 1	CCV63	19233	5.0980	-	7470A	PPB	08 Apr 2021 10:15:07	U	SMPL
CCB63 - 1	CCB63	-89	-0.0132	-	7470A	PPB	08 Apr 2021 10:17:22	U	SMPL
CRA - 1	CRA	615	0.1730	-	7470A	PPB	08 Apr 2021 10:19:37	U	SMPL
HighStd - 1	HighStd	37776	10.0032	-	7470A	PPB	08 Apr 2021 10:21:52	U	SMPL
ChkStd - 1	ChkStd	27941	7.4015	-	7470A	PPB	08 Apr 2021 10:24:07	U	SMPL
PB135610BL - 1	PBW	-215	-0.0466	-	7470A	PPB	08 Apr 2021 10:26:24	U	SMPL
M1854-09 - 1	TP-1S-(0-4)	24	0.0166	-	7470A	PPB	08 Apr 2021 10:30:58	U	SMPL
M1854-10 - 1	TP-1D-(4-10)	92	0.0346	-	7470A	PPB	08 Apr 2021 10:33:14	U	SMPL
M1888-09 - 1	TP-3S-(0-4)	115	0.0407	-	7470A	PPB	08 Apr 2021 10:35:30	U	SMPL
M1888-10 - 1	TP-3D-(4-10)	137	0.0465	-	7470A	PPB	08 Apr 2021 10:37:48	U	SMPL
M1889-05 - 1	TP-5-(0-4)	147	0.0492	-	7470A	PPB	08 Apr 2021 10:40:02	U	SMPL
CCV64 - 1	CCV64	17866	4.7364	-	7470A	PPB	08 Apr 2021 10:42:18	U	SMPL
CCB64 - 1	CCB64	-61	-0.0058	-	7470A	PPB	08 Apr 2021 10:44:33	U	SMPL
M1911-08 - 1	TP-6S-(0-4)	359	0.1053	-	7470A	PPB	08 Apr 2021 10:46:48	U	SMPL
M1911-09 - 1	TP-6D-(4-10)	131	0.0450	-	7470A	PPB	08 Apr 2021 10:49:04	U	SMPL
M1961-02 - 1	TR1-A-D-COMP	135	0.0460	-	7470A	PPB	08 Apr 2021 10:51:19	U	SMPL
M1961-13 - 1	TR2-A-D-COMP	155	0.0513	-	7470A	PPB	08 Apr 2021 10:53:35	U	SMPL
M1961-13DUP - 1	TR2-A-D-COMPDUP	116	0.0410	-	7470A	PPB	08 Apr 2021 10:55:50	U	SMPL
M1961-13MS - 1	TR2-A-D-COMPMS	16430	4.3565	-	7470A	PPB	08 Apr 2021 10:58:07	U	SMPL
M1961-13MSD - 1	TR2-A-D-COMPMSD	16180	4.2904	-	7470A	PPB	08 Apr 2021 11:00:23	U	SMPL
PB135611BL - 1	PBW	-30	0.0024	-	7470A	PPB	08 Apr 2021 11:02:39	U	SMPL
PB135611BS - 1	LCSW	13373	3.5478	-	7470A	PPB	08 Apr 2021 11:07:16	U	SMPL
M1946-01 - 1	FS-SB10	322	0.0955	-	7470A	PPB	08 Apr 2021 11:09:31	U	SMPL
CCV65 - 1	CCV65	19650	5.2083	-	7470A	PPB	08 Apr 2021 11:11:46	U	SMPL
CCB65 - 1	CCB65	-116	-0.0204	-	7470A	PPB	08 Apr 2021 11:14:01	U	SMPL
M1946-02 - 1	FS-SB10X	485	0.1386	-	7470A	PPB	08 Apr 2021 11:16:17	U	SMPL
M1946-03 - 1	FS-SB9	387	0.1127	-	7470A	PPB	08 Apr 2021 11:18:32	U	SMPL
M1946-03DUP - 1	FS-SB9DUP	352	0.1034	-	7470A	PPB	08 Apr 2021 11:20:47	U	SMPL
M1946-04 - 1	M1946-03MS	16320	4.3274	-	7470A	PPB	08 Apr 2021 11:23:02	U	SMPL
M1946-05 - 1	M1946-03MSD	16491	4.3726	-	7470A	PPB	08 Apr 2021 11:25:18	U	SMPL
M1946-06 - 1	FS-SB11	-46	-0.0019	-	7470A	PPB	08 Apr 2021 11:27:35	U	SMPL
M1952-03 - 1	30820	6505	1.7311	-	7470A	PPB	08 Apr 2021 11:29:51	U	SMPL
M1955-03 - 1	2885	308	0.0918	-	7470A	PPB	08 Apr 2021 11:32:07	U	SMPL
M1964-01 - 1	FS-SB16	177	0.0571	-	7470A	PPB	08 Apr 2021 11:34:26	U	SMPL
PB135610BS - 1	LCSW	14487	3.8425	-	7470A	PPB	08 Apr 2021 11:36:42	U	SMPL
CCV66 - 1	CCV66	19575	5.1885	-	7470A	PPB	08 Apr 2021 11:38:57	U	SMPL
CCB66 - 1	CCB66	78	0.0309	-	7470A	PPB	08 Apr 2021 11:41:13	U	SMPL
PB135610TB - 1	PB135582TB	135	0.0460	-	7470A	PPB	08 Apr 2021 11:43:28	U	SMPL
M1961-13LX5 - 1		174	0.0563	-	7470A	PPB	08 Apr 2021 11:45:44	U	SMPL
M1961-13A - 1		16117	4.2737	-	7470A	PPB	08 Apr 2021 11:48:00	U	SMPL
M1946-03LX5 - 1		13	0.0137	-	7470A	PPB	08 Apr 2021 11:50:15	U	SMPL
M1946-03A - 1		15977	4.2367	-	7470A	PPB	08 Apr 2021 11:52:31	U	SMPL
CCV67 - 1	CCV67	18760	4.9729	-	7470A	PPB	08 Apr 2021 11:54:47	U	SMPL
CCB67 - 1	CCB67	-52	-0.0035	-	7470A	PPB	08 Apr 2021 11:57:04	U	SMPL