



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113874

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/08/2021	08:25
2	CAL2	1	Water	NA	NA	20.3	6.99	04/08/2021	08:26
3	CAL3	1	Water	NA	NA	20.3	10.03	04/08/2021	08:27
4	ICV	1	Water	NA	NA	20.4	7.00	04/08/2021	08:30
5	CCV1	1	Water	NA	NA	20.3	2.01	04/08/2021	08:32
6	M1965-01	1	Water	NA	NA	20.6	5.47	04/08/2021	08:35
7	M1967-12	1	Water	NA	NA	20.1	8.13	04/08/2021	08:36
8	M1967-13	1	Water	NA	NA	20.2	6.11	04/08/2021	08:38
9	M1974-01	1	Water	NA	NA	20.5	6.04	04/08/2021	08:40
10	M1974-01DUP	1	Water	NA	NA	20.5	6.05	04/08/2021	08:41
11	CCV2	1	Water	NA	NA	20.4	12.02	04/08/2021	08:43

WORKLIST(Hardcopy Internal Chain)

2B113874

WorkList Name : ph m1961 WorkList ID : 147836 Department : Wet-Chemistry Date : 04-08-2021 08:17:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/21/2021	Water	M1965-01	pH	Cool 4 deg C	STER17	O13	JS9600	03/21/2021	9040C
04/06/2021	Water	M1967-12	pH	Cool 4 deg C	BEMS01	O41	EB_2021-04-06_SO	04/06/2021	9040C
04/06/2021	Water	M1967-13	pH	Cool 4 deg C	BEMS01	O41	EB_2021-04-06_SED	04/06/2021	9040C
04/07/2021	Water	M1974-01	pH	Cool 4 deg C	PSEG03	O31	FM648	04/07/2021	9040C

Date/Time 04/08/21 08:20
 Raw Sample Received by: JO (wet chem)
 Raw Sample Relinquished by: Amy (wet chem)

Date/Time 04/08/21 9:35
 Raw Sample Received by: Amy (wet chem)
 Raw Sample Relinquished by: JO (wet chem)