



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113875

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	04/08/2021	10:05
2	CAL2	1	Water	NA	NA	20.4	7.00	04/08/2021	10:06
3	CAL3	1	Water	NA	NA	20.4	10.03	04/08/2021	10:07
4	ICV	1	Water	NA	NA	20.4	6.99	04/08/2021	10:08
5	CCV1	1	Water	NA	NA	20.4	2.01	04/08/2021	10:10
6	M1967-01	1	Solid	20.03	20	21.0	6.87	04/08/2021	10:11
7	M1967-01DUP	1	Solid	20.04	20	21.1	6.88	04/08/2021	10:12
8	M1967-04	1	Solid	20.02	20	21.1	7.04	04/08/2021	10:14
9	M1967-05	1	Solid	20.03	20	20.7	7.37	04/08/2021	10:15
10	M1967-06	1	Solid	20.05	20	20.5	10.64	04/08/2021	10:20
11	M1967-07	1	Solid	20.04	20	20.3	10.82	04/08/2021	10:22
12	M1967-08	1	Solid	20.03	20	20.5	9.13	04/08/2021	10:25
13	M1967-09	1	Solid	20.04	20	20.4	10.80	04/08/2021	10:30
14	M1967-10	1	Solid	20.02	20	20.5	10.67	04/08/2021	10:32
15	M1967-11	1	Solid	20.01	20	20.3	10.81	04/08/2021	10:35
16	CCV2	1	Water	NA	NA	20.4	12.02	04/08/2021	10:36

2B 113875

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1967

WorkList ID : 147864

Department : Wet-Chemistry

Date : 04-08-2021 09:49:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/06/2021	Solid	M1967-01	PH	Cool 4 deg C	BEMS01	O41	SED-1A	04/06/2021	9045D
04/06/2021	Solid	M1967-04	PH	Cool 4 deg C	BEMS01	O41	SED-1B	04/06/2021	9045D
04/06/2021	Solid	M1967-05	PH	Cool 4 deg C	BEMS01	O41	SED-1C	04/06/2021	9045D
04/06/2021	Solid	M1967-06	PH	Cool 4 deg C	BEMS01	O41	SB2-A	04/06/2021	9045D
04/06/2021	Solid	M1967-07	PH	Cool 4 deg C	BEMS01	O41	SB2-B	04/06/2021	9045D
04/06/2021	Solid	M1967-08	PH	Cool 4 deg C	BEMS01	O41	SB2-C	04/06/2021	9045D
04/06/2021	Solid	M1967-09	PH	Cool 4 deg C	BEMS01	O41	SB1-A	04/06/2021	9045D
04/06/2021	Solid	M1967-10	PH	Cool 4 deg C	BEMS01	O41	SB1-B	04/06/2021	9045D
04/06/2021	Solid	M1967-11	PH	Cool 4 deg C	BEMS01	O41	SB1-C	04/06/2021	9045D

Date/Time 04/08/21 9:52
Raw Sample Received by: JP (wet-chem)
Raw Sample Relinquished by: AM (wet-chem)

Date/Time 04/08/21 11:30
Raw Sample Received by: AM (wet-chem)
Raw Sample Relinquished by: JP (wet-chem)