

LB113897


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37560



A= 0.0000e+000

B= 2.6518e-004

C= -5.3288e-003

Rho= 0.9999549

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.003	0.003	30	0.000	30					
0.2	0.200	0.178	-0.022	693	0.0 %	693					
2.5	2.500	2.515	0.015	9503	0.0 %	9503					
5.0	5.000	4.983	-0.017	18813	0.0 %	18813					
7.5	7.500	7.566	0.066	28552	0.0 %	28552					
10.0	10.000	9.955	-0.045	37560	0.0 %	37560					

9.00
-11
1
0
1
0

LB113897 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Apr 2021 09:52:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	30	-	0.00007470A	PPB		09 Apr 2021 09:55:57	S	Std
0.2 - 1	SD-2	693	-	0.20007470A	PPB		09 Apr 2021 10:00:45	S	Std
2.5 - 1	SD-5	9503	-	2.50007470A	PPB		09 Apr 2021 10:03:02	S	Std
5.0 - 1	SD	18813	-	5.00007470A	PPB		09 Apr 2021 10:05:19	S	Std
7.5 - 1	SD-5	28552	-	7.50007470A	PPB		09 Apr 2021 10:07:36	S	Std
10.0 - 1	SIV	37560	-	10.00007470A	PPB		09 Apr 2021 10:12:29	S	Std
ICV78 - 1	ICV78	15409	4.0808	-7470A	PPB		09 Apr 2021 10:16:26	U	SMPL
ICB78 - 1	ICB78	-234	-0.0674	-7470A	PPB		09 Apr 2021 10:18:41	U	SMPL
CCV71 - 1	CCV71	18632	4.9355	-7470A	PPB		09 Apr 2021 10:20:57	U	SMPL
CCB71 - 1	CCB71	-250	-0.0716	-7470A	PPB		09 Apr 2021 10:25:27	U	SMPL
CRA - 1	CRA	611	0.1567	-7470A	PPB		09 Apr 2021 10:27:42	U	SMPL
HighStd - 1	HighStd	38371	10.1698	-7470A	PPB		09 Apr 2021 10:29:58	U	SMPL
ChkStd - 1	ChkStd	25655	6.7978	-7470A	PPB		09 Apr 2021 10:32:15	U	SMPL
PB135636BL - 1	PBW	-327	-0.0920	-7470A	PPB		09 Apr 2021 10:34:32	U	SMPL
PB135636BS - 1	LCSW	15255	4.0400	-7470A	PPB		09 Apr 2021 10:36:48	U	SMPL
M1966-01 - 1	MW-122	9572	2.5330	-7470A	PPB		09 Apr 2021 10:39:06	U	SMPL
M1966-02 - 1	MW-119	165	0.0384	-7470A	PPB		09 Apr 2021 10:41:23	U	SMPL
M1966-03 - 1	MW-110RR	804	0.2079	-7470A	PPB		09 Apr 2021 10:43:41	U	SMPL
M1966-04 - 1	MW-115RR	741	0.1912	-7470A	PPB		09 Apr 2021 10:45:56	U	SMPL
M1966-05 - 1	MW-123R	375	0.0941	-7470A	PPB		09 Apr 2021 10:48:11	U	SMPL
CCV72 - 1	CCV72	18159	4.8100	-7470A	PPB		09 Apr 2021 10:50:26	U	SMPL
CCB72 - 1	CCB72	-79	-0.0263	-7470A	PPB		09 Apr 2021 10:55:30	U	SMPL
M1966-05DUP - 1	MW-123RDUP	288	0.0710	-7470A	PPB		09 Apr 2021 11:07:38	U	SMPL
M1966-06 - 1	M1966-05MS	12871	3.4078	-7470A	PPB		09 Apr 2021 11:12:29	U	SMPL
M1966-07 - 1	M1966-05MSD	13817	3.6586	-7470A	PPB		09 Apr 2021 11:14:45	U	SMPL
M1966-08 - 1	EB_2021-04-06_GW	-326	-0.0918	-7470A	PPB		09 Apr 2021 11:17:03	U	SMPL
M1967-12 - 1	EB_2021-04-06_SO	-54	-0.0196	-7470A	PPB		09 Apr 2021 11:19:19	U	SMPL
M1967-13 - 1	EB_2021-04-06_SED	-19	-0.0104	-7470A	PPB		09 Apr 2021 11:21:35	U	SMPL
M1974-01 - 1	FM648	259	0.0634	-7470A	PPB		09 Apr 2021 11:23:52	U	SMPL
M1985-03 - 1	FM119	2515	0.6616	-7470A	PPB		09 Apr 2021 11:26:09	U	SMPL
M1966-05LX5 - 1		143	0.0326	-7470A	PPB		09 Apr 2021 11:28:25	U	SMPL
M1966-05A - 1		15331	4.0601	-7470A	PPB		09 Apr 2021 11:30:40	U	SMPL
CCV73 - 1	CCV73	19032	5.0415	-7470A	PPB		09 Apr 2021 11:32:55	U	SMPL
CCB73 - 1	CCB73	-125	-0.0385	-7470A	PPB				