



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113899

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/09/2021	11:10
2	M1983-04		1	15	100.00	>212.0	>212.0	04/09/2021	11:40
3	M1983-04DUP		1	15	100.00	>212.0	>212.0	04/09/2021	12:05
4	M1983-06		1	12	100.00	>212.0	>212.0	04/09/2021	13:05
5	M1985-03		1	14	100.00	>212.0	>212.0	04/09/2021	13:40
6	M1990-01		1	13	100.00	>212.0	>212.0	04/09/2021	14:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB113899

WorkList Name : fp4921 WorkList ID : 147895 Department : Wet-Chemistry Date : 04-09-2021 09:00:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/15/2021	Water	M1983-04	Flash Point	Cool 4 deg C	PSEG03	O32	41120	04/08/2021	1010A
04/15/2021	Water	M1983-06	Flash Point	Cool 4 deg C	PSEG03	O32	31019	04/08/2021	1010A
04/13/2021	Water	M1985-03	Flash Point	Cool 4 deg C	PSEG03	O32	FM119	04/08/2021	1010A
04/15/2021	Water	M1990-01	Flash Point	Cool 4 deg C	TULL02		ELEVATOR-PIT-STANDIN	04/08/2021	1010A

Date/Time 04/09/21 10:40
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: SP (Wetchem)

Date/Time 04/09/21 14:10
 Raw Sample Received by: SP (Wetchem)
 Raw Sample Relinquished by: KS (Wetchem)