



TOTAL VOLATILE SOLIDS 160.4

TEMP1 IN: 104 °C 04/09/2021 16:00 TEMP1 OUT: 103 °C 04/12/2021 08:00
TEMP2 IN: 103 °C 04/12/2021 08:30 TEMP2 OUT: 103 °C 04/12/2021 10:15
TEMP3 IN: 540 °C 04/12/2021 10:40 TEMP3 OUT: 540 °C 04/12/2021 12:15
TEMP4 IN: 560 °C 04/12/2021 12:45 TEMP4 OUT: 560 °C 04/12/2021 14:00

Run Number: LB113905

SUPERVISOR: apatel

ANALYST: jignesh

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Empty Dish + Sample Weight (g)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (%)
1	LB113905BL	65.7841	65.7841	65.7841	65.7842	65.7842	65.7842	65.7842	0.0000	0
2	M1991-01	59.5623	59.5623	79.9721	78.6193	78.6193	78.1126	78.1126	0.5067	2.7
3	M1991-05	58.3178	58.3178	79.0557	77.5114	77.5114	76.9238	76.9238	0.5876	3.1
4	M1991-09	57.1841	57.1841	78.3829	76.2244	76.2244	75.5169	75.5169	0.7075	3.7
5	M1991-13	55.8885	55.8885	78.9759	77.5472	77.5472	76.9979	76.9979	0.5493	2.5
6	M1991-15	59.9797	59.9797	80.2887	78.5090	78.509	77.8180	77.8180	0.6910	3.7
7	M1991-17	59.3941	59.3941	80.2932	78.4997	78.4997	77.8484	77.8484	0.6513	3.4
8	M1991-17DUP	53.5513	53.5513	74.2692	72.4304	72.4304	71.6605	71.6605	0.7699	4.1

A = Sample Weight (g)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{F - E} \star 100$

WORKLIST(Hardcopy Internal Chain)

VB 112905

WorkList Name : TVS M1991 WorkList ID : 147904 Department : Wet-Chemistry Date : 04-09-2021 12:16:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/13/2021	Solid	M1991-01	TVS	Cool 4 deg C	ENTA05	O21	JCSY-WC1-1	04/08/2021	160.4
04/13/2021	Solid	M1991-05	TVS	Cool 4 deg C	ENTA05	O21	JCSY-WC2-1	04/08/2021	160.4
04/13/2021	Solid	M1991-09	TVS	Cool 4 deg C	ENTA05	O21	JCSY-WC3-1	04/08/2021	160.4
04/13/2021	Solid	M1991-13	TVS	Cool 4 deg C	ENTA05	O21	JCSY-WC1-A	04/08/2021	160.4
04/13/2021	Solid	M1991-15	TVS	Cool 4 deg C	ENTA05	O21	JCSY-WC2-A	04/08/2021	160.4
04/13/2021	Solid	M1991-17	TVS	Cool 4 deg C	ENTA05	O21	JCSY-WC3-A	04/08/2021	160.4

Date/Time 04/09/21 12:30
 Raw Sample Received by: JB Chaitanya
 Raw Sample Relinquished by: JB Chaitanya

Date/Time 04/09/21 16:30
 Raw Sample Received by: JB Chaitanya
 Raw Sample Relinquished by: JB Chaitanya