



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113909

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/09/2021	12:35
2	CAL2	1	Water	NA	NA	20.3	7.00	04/09/2021	12:36
3	CAL3	1	Water	NA	NA	20.3	10.03	04/09/2021	12:37
4	ICV	1	Water	NA	NA	20.3	6.99	04/09/2021	12:39
5	CCV1	1	Water	NA	NA	20.3	2.01	04/09/2021	12:40
6	M1956-02	1	Solid	20.02	20	20.3	5.56	04/09/2021	12:41
7	M1991-02	1	Solid	20.02	20	20.9	9.06	04/09/2021	12:42
8	M1991-02DUP	1	Solid	20.03	20	20.8	9.07	04/09/2021	12:43
9	CCV2	1	Water	NA	NA	20.3	12.02	04/09/2021	12:45

113909

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M1991 WorkList ID : 147909 Department : Wet-Chemistry Date : 04-09-2021 12:16:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/06/2021	Solid	M1956-02	Corrosivity	Cool 4 deg C	PSEG03	O23	OIL-DEBRIS-3-5	04/06/2021	9045D
04/08/2021	Solid	M1991-02	Corrosivity	Cool 4 deg C	ENTA05	O21	JCSY-WC1-1	04/08/2021	9045D

Date/Time 04/09/21 12:20
 Raw Sample Received by: JQ (wet Chem)
 Raw Sample Relinquished by: JQ (wet Chem)

Date/Time 04/09/21 16:30
 Raw Sample Received by: JQ (wet Chem)
 Raw Sample Relinquished by: JQ (wet Chem)