

LB118910


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44603



A= 0.0000e+000

B= 2.2463e-004 Slope

C= -4.1992e-003 y-intercept

Rho= 0.9999284

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.009	0.009	58	0.000	58					1
0.05	0.050					204					-5
0.20	0.200	0.190	-0.010	866	0.0 %	866					-1
2.50	2.500	2.471	-0.029	11019	0.0 %	11019					2
5.00	5.000	5.079	0.079	22631	0.0 %	22631					-1
7.50	7.500	7.436	-0.064	33120	0.0 %	33120					0
10.0	10.000	10.015	0.015	44603	0.0 %	44603					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	58	-		0 ISM02.4	PPB	4/9/2021 13:39	Std
	0.2 S0.2	866	-		0.2 ISM02.4	PPB	4/9/2021 13:41	Std
	2.5 S2.5	11019	-		2.5 ISM02.4	PPB	4/9/2021 13:43	Std
	5 S5.0	22631	-		5 ISM02.4	PPB	4/9/2021 13:46	Std
	7.5 S7.5	33120	-		7.5 ISM02.4	PPB	4/9/2021 13:48	Std
	10 S10.0	44603	-		10 ISM02.4	PPB	4/9/2021 13:53	Std
ICV	ICV	16158	3.6254 -		ISM02.4	PPB	4/9/2021 13:55	SMPL
ICB	ICB	-172	-0.0428 -		ISM02.4	PPB	4/9/2021 13:58	SMPL
CCV074	CCV074	22826	5.1232 -		ISM02.4	PPB	4/9/2021 14:00	SMPL
CCB074	CCB074	-169	-0.0422 -		ISM02.4	PPB	4/9/2021 14:02	SMPL
PB135645BL	PBW645	7	-0.0026 -		ISM02.4	PPB	4/9/2021 14:04	SMPL
M1930-02	MESQF7	435	0.0935 -		ISM02.4	PPB	4/9/2021 14:07	SMPL
M1930-03	MESQF7D	492	0.1063 -		ISM02.4	PPB	4/9/2021 14:11	SMPL
M1930-04	MESQF7S	5071	1.1349 -		ISM02.4	PPB	4/9/2021 14:16	SMPL
M1930-12	MESQG4	-18	-0.0082 -		ISM02.4	PPB	4/9/2021 14:19	SMPL
PB135646BL	PBW646	21	0.0005 -		ISM02.4	PPB	4/9/2021 14:21	SMPL
M1943-01	MYARR4	161	0.032 -		ISM02.4	PPB	4/9/2021 14:23	SMPL
M1943-02	MYARR9	98	0.0178 -		ISM02.4	PPB	4/9/2021 14:25	SMPL
M1943-03	MYARR9D	84	0.0147 -		ISM02.4	PPB	4/9/2021 14:28	SMPL
M1943-04	MYARR9S	4564	1.021 -		ISM02.4	PPB	4/9/2021 14:30	SMPL
PB135647BL	PBW647	-5	-0.0053 -		ISM02.4	PPB	4/9/2021 14:32	SMPL
M1944-01	MYARR5	153	0.0302 -		ISM02.4	PPB	4/9/2021 14:34	SMPL
M1944-02	MYARR6	117	0.0221 -		ISM02.4	PPB	4/9/2021 14:37	SMPL
M1944-03	MYARR6D	118	0.0223 -		ISM02.4	PPB	4/9/2021 14:39	SMPL
M1944-04	MYARR6S	4480	1.0021 -		ISM02.4	PPB	4/9/2021 14:41	SMPL
M1944-05	MYARR8	176690	39.6855 -		ISM02.4	PPB	4/9/2021 14:43	SMPL
M1944-06	MYARS0	-8	-0.006 -		ISM02.4	PPB	4/9/2021 14:48	SMPL
M1944-05DLX10	MYARR8	20417	4.5821 -		ISM02.4	PPB	4/9/2021 14:51	SMPL
CCV075	CCV075	22487	5.047 -		ISM02.4	PPB	4/9/2021 14:53	SMPL
CCB075	CCB075	-9	-0.0062 -		ISM02.4	PPB	4/9/2021 14:55	SMPL