

LB118913


7471B

INSTRUMENT ID: CV4

Linear

 μ Abs.:

55428



A= 0.0000e+000

B= 1.7923e-004

C= 1.9141e-002

Rho= 0.9999366

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.029	0.029	53	0.000	53					-10
0.2	0.200	0.181	-0.019	903	0.0 %	903					-2
2.5	2.500	2.448	-0.052	13554	0.0 %	13554					1
5.0	5.000	5.030	0.030	27960	0.0 %	27960					1
7.5	7.500	7.558	0.058	42066	0.0 %	42066					0
10.0	10.000	9.953	-0.047	55428	0.0 %	55428					

LB113913 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 09 Apr 2021 15:20:13					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	53	-	0.0000	7471B	PPB	09 Apr 2021 15:21:55	S	Std
0.2 - 1	SP 2	903	-	0.2000	7471B	PPB	09 Apr 2021 15:24:12	S	Std
2.5 - 1	52.5	13554	-	2.5000	7471B	PPB	09 Apr 2021 15:26:29	S	Std
5.0 - 1	55	27960	-	5.0000	7471B	PPB	09 Apr 2021 15:28:46	S	Std
7.5 - 1	57.5	42066	-	7.5000	7471B	PPB	09 Apr 2021 15:31:02	S	Std
10.0 - 1	510	55428	-	10.0000	7471B	PPB	09 Apr 2021 15:35:58	S	Std
ICV80 - 1	ICV80	22725	4.0920	-7471B	PPB	PPB	09 Apr 2021 15:38:35	U	SMPL
ICB80 - 1	ICB80	-118	-0.0020	-7471B	PPB	PPB	09 Apr 2021 15:40:49	U	SMPL
CCV76 - 1	CCV76	28510	5.1289	-7471B	PPB	PPB	09 Apr 2021 15:43:05	U	SMPL
CCB76 - 1	CCB76	-115	-0.0015	-7471B	PPB	PPB	09 Apr 2021 15:45:20	U	SMPL
CRA - 1	CRA	876	0.1761	-7471B	PPB	PPB	09 Apr 2021 15:47:36	U	SMPL
HighStd - 1	HighStd	55152	9.9038	-7471B	PPB	PPB	09 Apr 2021 15:49:50	U	SMPL
ChkStd - 1	ChkStd	39549	7.1073	-7471B	PPB	PPB	09 Apr 2021 15:52:05	U	SMPL
PB135637BL - 1	PB135637BL	-280	-0.0310	-7471B	PPB	PPB	09 Apr 2021 15:54:22	U	SMPL
PB135637BS - 1	PB135637BS	20300	3.6574	-7471B	PPB	PPB	09 Apr 2021 15:56:40	U	SMPL
M1947-01 - 1	FS-SB19B(6-7)	2206	0.4145	-7471B	PPB	PPB	09 Apr 2021 15:58:56	U	SMPL
M1947-02 - 1	FS-SB19B(19-20)	1439	0.2770	-7471B	PPB	PPB	09 Apr 2021 16:01:12	U	SMPL
M1947-03 - 1	FS-SB19B(19-20)X	2178	0.4095	-7471B	PPB	PPB	09 Apr 2021 16:03:29	U	SMPL
M1947-07 - 1	FS-SB9(20-21)	9786	1.7730	-7471B	PPB	PPB	09 Apr 2021 16:05:45	U	SMPL
M1967-01 - 1	SED-1A	20111	3.6236	-7471B	PPB	PPB	09 Apr 2021 16:08:01	U	SMPL
CCV77 - 1	CCV77	28284	5.0884	-7471B	PPB	PPB	09 Apr 2021 16:10:16	U	SMPL
CCB77 - 1	CCB77	-84	0.0041	-7471B	PPB	PPB	09 Apr 2021 16:12:30	U	SMPL
M1967-01DUP - 1	SED-1ADUP	19862	3.5789	-7471B	PPB	PPB	09 Apr 2021 16:14:45	U	SMPL
M1967-02 - 1	M1967-01MS	39416	7.0835	-7471B	PPB	PPB	09 Apr 2021 16:17:00	U	SMPL
M1967-03 - 1	M1967-01MSD	42344	7.6083	-7471B	PPB	PPB	09 Apr 2021 16:19:16	U	SMPL
M1967-04 - 1	SED-1B	46227	8.3042	-7471B	PPB	PPB	09 Apr 2021 16:21:32	U	SMPL
M1967-05 - 1	SED-1C	13239	2.3919	-7471B	PPB	PPB	09 Apr 2021 16:23:49	U	SMPL
M1967-06 - 1	SB2-A	50344	9.0421	-7471B	PPB	PPB	09 Apr 2021 16:26:06	U	SMPL
M1967-07 - 1	SB2-B	49519	8.8942	-7471B	PPB	PPB	09 Apr 2021 16:28:23	U	SMPL
M1967-08 - 1	SB2-C	9822	1.7795	-7471B	PPB	PPB	09 Apr 2021 16:30:41	U	SMPL
M1967-09 - 1	SB1-A	312839	56.0879	-7471B	PPB	PPB	09 Apr 2021 16:32:59	U	SMPL
M1967-10 - 1	SB1-B	67674	12.1481	-7471B	PPB	PPB	09 Apr 2021 16:35:14	U	SMPL
CCV78 - 1	CCV78	28914	5.2013	-7471B	PPB	PPB	09 Apr 2021 16:37:33	U	SMPL
CCB78 - 1	CCB78	-115	-0.0015	-7471B	PPB	PPB	09 Apr 2021 16:39:50	U	SMPL
M1967-11 - 1	SB1-C	50532	9.0758	-7471B	PPB	PPB	09 Apr 2021 16:42:05	U	SMPL
M1983-01 - 1	41018	1951	0.3688	-7471B	PPB	PPB	09 Apr 2021 16:44:20	U	SMPL
M1983-02 - 1	16873	714	0.1471	-7471B	PPB	PPB	09 Apr 2021 16:46:37	U	SMPL
M1983-03 - 1	41019	2080	0.3919	-7471B	PPB	PPB	09 Apr 2021 16:48:53	U	SMPL
M1985-01 - 1	72-11939-72-11936	4094	0.7529	-7471B	PPB	PPB	09 Apr 2021 16:51:08	U	SMPL
M1967-01LX5 - 1		3866	0.7120	-7471B	PPB	PPB	09 Apr 2021 16:53:24	U	SMPL
M1967-01A - 1		41526	7.4617	-7471B	PPB	PPB	09 Apr 2021 16:55:41	U	SMPL
M1967-09DLX10 - 1		36878	6.6286	-7471B	PPB	PPB	09 Apr 2021 16:57:57	U	SMPL
M1967-10DLX2 - 1		44476	7.9904	-7471B	PPB	PPB	09 Apr 2021 17:00:56	U	SMPL
CCV79 - 1	CCV79	28712	5.1651	-7471B	PPB	PPB	09 Apr 2021 17:03:11	U	SMPL
CCB79 - 1	CCB79	-69	0.0068	-7471B	PPB	PPB	09 Apr 2021 17:05:27	U	SMPL