



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113921

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/09/2021	17:00
2	CAL2	1	Water	NA	NA	20.3	7.00	04/09/2021	17:01
3	CAL3	1	Water	NA	NA	20.3	10.03	04/09/2021	17:02
4	ICV	1	Water	NA	NA	20.3	7.02	04/09/2021	17:05
5	CCV1	1	Water	NA	NA	20.3	2.01	04/09/2021	17:06
6	M1994-01	1	Water	NA	NA	20.1	6.39	04/09/2021	17:08
7	M1995-01	1	Water	NA	NA	20.6	7.33	04/09/2021	17:10
8	M1998-13	1	Water	NA	NA	20.2	5.38	04/09/2021	17:11
9	M1998-13DUP	1	Water	NA	NA	20.3	5.39	04/09/2021	17:12
10	CCV2	1	Water	NA	NA	20.2	12.02	04/09/2021	17:15

WORKLIST(Hardcopy Internal Chain)

18913021

WorkList Name : PH M1994 WorkList ID : 147913 Department : Wet-Chemistry Date : 04-09-2021 16:49:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/09/2021	Water	M1994-01	C pH	Cool 4 deg C	PSEG03	O33	TRE-1170	04/09/2021	9040C
04/09/2021	Water	M1995-01	C pH	Cool 4 deg C	PSEG03	O33	NWB-1592	04/09/2021	9040C
04/08/2021	Water	M1998-13	E pH	Cool 4 deg C	BEMS01	O22	EB_2021-04-08-2021_SC	04/08/2021	9040C

Date/Time 04/09/21 16:50
Raw Sample Received by: JP (wet chem)
Raw Sample Relinquished by: KS (wet chem)

Date/Time 04/09/21 17:50
Raw Sample Received by: KS (wet chem)
Raw Sample Relinquished by: JP (wet chem)