



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113922

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/09/2021	17:00
2	CAL2	1	Water	NA	NA	20.3	7.00	04/09/2021	17:01
3	CAL3	1	Water	NA	NA	20.3	10.03	04/09/2021	17:02
4	ICV	1	Water	NA	NA	20.3	7.02	04/09/2021	17:05
5	CCV1	1	Water	NA	NA	20.3	2.01	04/09/2021	17:06
6	M1992-01	1	Solid	20.03	20	21.4	9.16	04/09/2021	17:20
7	M1992-01DUP	1	Solid	20.04	20	21.5	9.17	04/09/2021	17:21
8	CCV2	1	Water	NA	NA	20.5	12.02	04/09/2021	17:22

113922

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1992 S WorkList ID : 147914 Department : Wet-Chemistry Date : 04-09-2021 16:49:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/09/2021	Solid	M1992-01	pH	Cool 4 deg C	PSEG05	O33	AU-05-040921	04/09/2021	9045D

Date/Time 04/09/21 16:50
 Raw Sample Received by: JP (wetchem)
 Raw Sample Relinquished by: KS (wetchem)

Date/Time 04/09/21 17:50
 Raw Sample Received by: KS (wetchem)
 Raw Sample Relinquished by: JP (wetchem) Chlen