



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113937

Slope : 97.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/12/2021	16:10
2	CAL2	1	Water	NA	NA	20.3	7.00	04/12/2021	16:11
3	CAL3	1	Water	NA	NA	20.3	10.03	04/12/2021	16:12
4	ICV	1	Water	NA	NA	20.4	6.99	04/12/2021	16:15
5	CCV1	1	Water	NA	NA	20.3	2.01	04/12/2021	16:16
6	M1982-03	1	Solid	20.02	20	21.4	8.26	04/12/2021	16:30
7	M1982-03DUP	1	Solid	20.01	20	21.4	8.27	04/12/2021	16:31
8	M1991-06	1	Solid	20.03	20	20.9	9.54	04/12/2021	16:35
9	M1991-10	1	Solid	20.04	20	20.7	8.60	04/12/2021	16:36
10	M2008-03	1	Solid	20.04	20	21.9	7.62	04/12/2021	16:40
11	M2016-01	1	Solid	20.02	20	20.3	5.14	04/12/2021	16:42
12	M2017-02	1	Solid	20.03	20	21.6	6.82	04/12/2021	16:44
13	M2019-02	1	Solid	20.05	20	21.4	9.32	04/12/2021	16:46
14	CCV2	1	Water	NA	NA	20.4	12.02	04/12/2021	16:48

WORKLIST(Hardcopy Internal Chain)

2013931

WorkList Name : CORROSIONITY M2014 WorkList ID : 147957 Department : Wet-Chemistry Date : 04-12-2021 15:49:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/12/2021	Solid	M1982-03	Corrosivity	Cool 4 deg C	DIST01	O51	WC-1	04/12/2021	9045D
04/08/2021	Solid	M1991-06	Corrosivity	Cool 4 deg C	ENTA05		JCSY-WC2-1	04/08/2021	9045D
04/08/2021	Solid	M1991-10	Corrosivity	Cool 4 deg C	ENTA05		JCSY-WC3-1	04/08/2021	9045D
04/12/2021	Solid	M2008-03	Corrosivity	Cool 4 deg C	DIST01	O13	NANUET-ST-041221	04/12/2021	9045D
04/12/2021	Solid	M2016-01	Corrosivity	Cool 4 deg C	PSEG03	O41	UNLABELED-24OZ-JAR	04/12/2021	9045D
04/12/2021	Solid	M2017-02	Corrosivity	Cool 4 deg C	FRAN01	O32	FISHKILL-SP-1	04/12/2021	9045D
04/12/2021	Solid	M2019-02	Corrosivity	Cool 4 deg C	TULL02	O41	149TH-ST-E1-101	04/12/2021	9045D

Date/Time 04/12/2021 16:10
 Raw Sample Received by: JD (wet chem)
 Raw Sample Relinquished by: KS (wet chem)

Date/Time 04-12-2021 17:30
 Raw Sample Received by: KS (wet chem)
 Raw Sample Relinquished by: JD (wet chem)