



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113939

Slope : 97.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/12/2021	16:10
2	CAL2	1	Water	NA	NA	20.3	7.00	04/12/2021	16:11
3	CAL3	1	Water	NA	NA	20.3	10.03	04/12/2021	16:12
4	ICV	1	Water	NA	NA	20.4	6.99	04/12/2021	16:15
5	CCV1	1	Water	NA	NA	20.3	2.01	04/12/2021	16:16
6	M2014-04	1	Water	NA	NA	20.6	11.12	04/12/2021	16:19
7	M2014-04DUP	1	Water	NA	NA	20.7	11.13	04/12/2021	16:20
8	CCV2	1	Water	NA	NA	20.3	12.02	04/12/2021	16:22

1B 113939

WORKLIST(Hardcopy Internal Chain)

Date : 04-12-2021 15:50:38

Department : Wet-Chemistry

WorkList ID : 147958

WorkList Name : CORROSIVITY M2015 V

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/12/2021	Water	M2014-04	Corrosivity	Cool 4 deg C	PSEG03		16128	04/12/2021	9040C

Date/Time 04.12.2021 17:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 04.12.2021 15:50
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]