


LB118944

7470A

INSTRUMENT ID: CV4

 μ Abs.:

37314



Linear

A= 0.0000e+000

B= 2.7106e-004

C= -7.3689e-002

Rho= 0.9998953

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.023	-0.023	186	0.000	186					3
0.2	0.200	0.205	0.005	1028	0.0 %	1028					1
2.5	2.500	2.521	0.021	9574	0.0 %	9574					1
5.0	5.000	5.060	0.060	18939	0.0 %	18939					1
7.5	7.500	7.396	-0.104	27557	0.0 %	27557					1
10.0	10.000	10.041	0.041	37314	0.0 %	37314					0

LB113944 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Apr 2021 09:44:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	186	-	0.0000	7470A	PPB	13 Apr 2021 09:49:14	S	Std
0.2 - 1	50.2	1028	-	0.2000	7470A	PPB	13 Apr 2021 09:51:31	S	Std
2.5 - 1	52.5	9574	-	2.5000	7470A	PPB	13 Apr 2021 09:53:48	S	Std
5.0 - 1	55	18939	-	5.0000	7470A	PPB	13 Apr 2021 09:56:05	S	Std
7.5 - 1	57.5	27557	-	7.5000	7470A	PPB	13 Apr 2021 09:58:22	S	Std
10.0 - 1	510	37314	-	10.0000	7470A	PPB	13 Apr 2021 10:05:52	S	Std
ICV82 - 1	ICV82	15756	4.1972	-	7470A	PPB	13 Apr 2021 10:08:34	U	SMPL
ICB82 - 1	ICB82	159	-0.0306	-	7470A	PPB	13 Apr 2021 10:13:22	U	SMPL
CCV86 - 1	CCV86	18555	4.9559	-	7470A	PPB	13 Apr 2021 10:15:37	U	SMPL
CCB86 - 1	CCB86	66	-0.0558	-	7470A	PPB	13 Apr 2021 10:17:52	U	SMPL
CRA - 1	CRA	984	0.1930	-	7470A	PPB	13 Apr 2021 10:20:07	U	SMPL
HighStd - 1	HighStd	38454	10.3498	-	7470A	PPB	13 Apr 2021 10:22:21	U	SMPL
ChkStd - 1	ChkStd	24240	6.4969	-	7470A	PPB	13 Apr 2021 10:24:36	U	SMPL
PB135679BL - 1	PBW	145	-0.0344	-	7470A	PPB	13 Apr 2021 10:26:54	U	SMPL
PB135679BS - 1	LCSW	14018	3.7261	-	7470A	PPB	13 Apr 2021 10:29:11	U	SMPL
M1991-02 - 1	JCSY-WC1-1	102	-0.0460	-	7470A	PPB	13 Apr 2021 10:31:27	U	SMPL
M1991-06 - 1	JCSY-WC2-1	324	0.0141	-	7470A	PPB	13 Apr 2021 10:33:45	U	SMPL
M1991-10 - 1	JCSY-WC3-1	363	0.0247	-	7470A	PPB	13 Apr 2021 10:36:02	U	SMPL
M1991-10DUP - 1	JCSY-WC3-1DUP	546	0.0743	-	7470A	PPB	13 Apr 2021 10:38:19	U	SMPL
M1991-10MS - 1	JCSY-WC3-1MS	14462	3.8464	-	7470A	PPB	13 Apr 2021 10:40:34	U	SMPL
CCV87 - 1	CCV87	19980	5.3422	-	7470A	PPB	13 Apr 2021 10:42:50	U	SMPL
CCB87 - 1	CCB87	72	-0.0542	-	7470A	PPB	13 Apr 2021 10:45:05	U	SMPL
M1991-10MSD - 1	JCSY-WC3-1MSD	14444	3.8416	-	7470A	PPB	13 Apr 2021 10:47:21	U	SMPL
PB135680BL - 1	PBW	114	-0.0428	-	7470A	PPB	13 Apr 2021 10:49:37	U	SMPL
PB135680BS - 1	LCSW	13988	3.7180	-	7470A	PPB	13 Apr 2021 10:51:52	U	SMPL
M1994-01 - 1	TRE-1170	317	0.0122	-	7470A	PPB	13 Apr 2021 10:54:07	U	SMPL
M1994-01DUP - 1	TRE-1170DUP	394	0.0331	-	7470A	PPB	13 Apr 2021 10:56:22	U	SMPL
M1994-01MS - 1	TRE-1170MS	14634	3.8931	-	7470A	PPB	13 Apr 2021 10:58:39	U	SMPL
M1994-01MSD - 1	TRE-1170MSD	15018	3.9972	-	7470A	PPB	13 Apr 2021 11:00:55	U	SMPL
PB135681BL - 1	PBW	337	0.0177	-	7470A	PPB	13 Apr 2021 11:03:12	U	SMPL
PB135681BS - 1	LCSW	14481	3.8516	-	7470A	PPB	13 Apr 2021 11:05:29	U	SMPL
M1995-01 - 1	NWB-1592	176	-0.0260	-	7470A	PPB	13 Apr 2021 11:07:44	U	SMPL
CCV88 - 1	CCV88	19954	5.3351	-	7470A	PPB	13 Apr 2021 11:10:01	U	SMPL
CCB88 - 1	CCB88	71	-0.0544	-	7470A	PPB	13 Apr 2021 11:12:16	U	SMPL
M1995-01DUP - 1	NWB-1592DUP	421	0.0404	-	7470A	PPB	13 Apr 2021 11:14:33	U	SMPL
M1995-01MS - 1	NWB-1592MS	12354	3.2750	-	7470A	PPB	13 Apr 2021 11:16:48	U	SMPL
M1995-01MSD - 1	NWB-1592MSD	12965	3.4407	-	7470A	PPB	13 Apr 2021 11:19:03	U	SMPL
M1998-13 - 1	EB_04-08-2021_SO	123	-0.0403	-	7470A	PPB	13 Apr 2021 11:21:21	U	SMPL
PB135679TB - 1	PB135652TB	345	0.0198	-	7470A	PPB	13 Apr 2021 11:23:38	U	SMPL
PB135680TB - 1	PB135671TB	117	-0.0420	-	7470A	PPB	13 Apr 2021 11:25:54	U	SMPL
CCV89 - 1	CCV89	18596	4.9670	-	7470A	PPB	13 Apr 2021 11:33:05	U	SMPL
CCB89 - 1	CCB89	169	-0.0279	-	7470A	PPB	13 Apr 2021 11:35:22	U	SMPL
M1991-10LX5 - 1		192	-0.0216	-	7470A	PPB	13 Apr 2021 11:37:39	U	SMPL
M1991-10A - 1		14685	3.9069	-	7470A	PPB	13 Apr 2021 11:39:54	U	SMPL
M1994-01LX5 - 1		77	-0.0528	-	7470A	PPB	13 Apr 2021 11:42:10	U	SMPL
M1994-01A - 1		15127	4.0267	-	7470A	PPB	13 Apr 2021 11:44:25	U	SMPL
M1995-01LX5 - 1		9	-0.0712	-	7470A	PPB	13 Apr 2021 11:46:41	U	SMPL
M1995-01A - 1		14972	3.9847	-	7470A	PPB	13 Apr 2021 12:00:33	U	SMPL
CCV90 - 1	CCV90	19792	5.2912	-	7470A	PPB	13 Apr 2021 12:02:49	U	SMPL
CCB90 - 1	CCB90	100	-0.0466	-	7470A	PPB	13 Apr 2021 12:05:05	U	SMPL