



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113946

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.00	04/13/2021	11:05
2	CAL2	1	Water	NA	NA	20.2	7.00	04/13/2021	11:06
3	CAL3	1	Water	NA	NA	20.2	10.03	04/13/2021	11:07
4	ICV	1	Water	NA	NA	20.4	6.99	04/13/2021	11:08
5	CCV1	1	Water	NA	NA	20.2	2.01	04/13/2021	11:10
6	M2021-01	1	Water	NA	NA	20.1	6.68	04/13/2021	11:12
7	M2021-01DUP	1	Water	NA	NA	20.2	6.69	04/13/2021	11:20
8	M2021-03	1	Water	NA	NA	20.2	7.01	04/13/2021	11:22
9	CCV2	1	Water	NA	NA	20.4	12.02	04/13/2021	11:24

WORKLIST(Hardcopy Internal Chain)

AB 113946

WorkList Name : PH M2021

WorkList ID : 147984

Department : Wet-Chemistry

Date : 04-13-2021 10:55:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/12/2021	Water	M2021-01	✓ pH	Cool 4 deg C	APT101	O32	WGGM-21-WW-TSF-015	04/12/2021	SM 4500-H B
04/12/2021	Water	M2021-03	✓ pH	Cool 4 deg C	APT101	O32	WGGM-21-WW-TSF-016	04/12/2021	SM 4500-H B

Date/Time 04/13/21 11:00
 Raw Sample Received by: JD (wet chem)
 Raw Sample Relinquished by: CP/SM

Date/Time 04/13/2021 11:40
 Raw Sample Received by: CP/SM
 Raw Sample Relinquished by: JD (wet chem)