

LB113956


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52537



A= 0.0000e+000

B= 1.8850e-004

C= -9.9936e-003

Rho= 0.9997173

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.030	0.030	212	0.000	212					-3
0.2	0.200	0.195	-0.005	1088	0.0 %	1088					-4
2.5	2.500	2.388	-0.112	12721	0.0 %	12721					1
5.0	5.000	5.060	0.060	26898	0.0 %	26898					2
7.5	7.500	7.633	0.133	40546	0.0 %	40546					-1
10.0	10.000	9.893	-0.107	52537	0.0 %	52537					

LB113956 INSTRUMENT ID ; CV1

Method: 7471B		Operator: Admin		Date of Analysis: 13 Apr 2021 13:49:49					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	212	-	0.0000	7471B	PPB	13 Apr 2021 13:53:52	S	Std
0.2 - 1	50.2	1088	-	0.2000	7471B	PPB	13 Apr 2021 13:56:09	S	Std
2.5 - 1	52.5	12721	-	2.5000	7471B	PPB	13 Apr 2021 13:58:26	S	Std
5.0 - 1	55	26898	-	5.0000	7471B	PPB	13 Apr 2021 14:00:43	S	Std
7.5 - 1	57.5	40546	-	7.5000	7471B	PPB	13 Apr 2021 14:03:00	S	Std
10.0 - 1	510	52537	-	10.0000	7471B	PPB	13 Apr 2021 14:07:56	S	Std
ICV83 - 1	ICV83	21051	3.9582	-	7471B	PPB	13 Apr 2021 14:10:50	U	SMPL
ICB83 - 1	ICB83	89	0.0068	-	7471B	PPB	13 Apr 2021 14:13:05	U	SMPL
CCV93 - 1	CCV93	27570	5.1871	-	7471B	PPB	13 Apr 2021 14:15:21	U	SMPL
CCB93 - 1	CCB93	64	0.0021	-	7471B	PPB	13 Apr 2021 14:17:36	U	SMPL
CRA - 1	CRA	1069	0.1915	-	7471B	PPB	13 Apr 2021 14:19:52	U	SMPL
HighStd - 1	HighStd	53626	10.0987	-	7471B	PPB	13 Apr 2021 14:24:48	U	SMPL
ChkStd - 1	ChkStd	37274	7.0163	-	7471B	PPB	13 Apr 2021 14:27:03	U	SMPL
PB135710BL - 1	PBS	-201	-0.0479	-	7471B	PPB	13 Apr 2021 14:29:20	U	SMPL
PB135710BS - 1	LCSS	18930	3.5584	-	7471B	PPB	13 Apr 2021 14:31:37	U	SMPL
M1982-02 - 1	WC-1	5968	1.1150	-	7471B	PPB	13 Apr 2021 14:33:53	U	SMPL
M2008-02 - 1	NANUET-ST-041221	11304	2.1209	-	7471B	PPB	13 Apr 2021 14:36:10	U	SMPL
M2009-01 - 1	OR-02-041221	2232	0.4107	-	7471B	PPB	13 Apr 2021 14:38:27	U	SMPL
M2010-01 - 1	CG-SP	969	0.1727	-	7471B	PPB	13 Apr 2021 14:40:44	U	SMPL
M2015-01 - 1	CL-02-041221	7714	1.4441	-	7471B	PPB	13 Apr 2021 14:42:59	U	SMPL
CCV94 - 1	CCV94	27420	5.1588	-	7471B	PPB	13 Apr 2021 14:45:14	U	SMPL
CCB94 - 1	CCB94	151	0.0185	-	7471B	PPB	13 Apr 2021 14:47:29	U	SMPL
M2017-01 - 1	FISHKILL-SP-1	5508	1.0283	-	7471B	PPB	13 Apr 2021 14:49:44	U	SMPL
M2017-01DUP - 1	FISHKILL-SP-1DUP	5491	1.0251	-	7471B	PPB	13 Apr 2021 14:52:00	U	SMPL
M2017-01MS - 1	FISHKILL-SP-1MS	26646	5.0129	-	7471B	PPB	13 Apr 2021 14:54:15	U	SMPL
M2017-01MSD - 1	FISHKILL-SP-1MSD	25914	4.8749	-	7471B	PPB	13 Apr 2021 14:56:31	U	SMPL
M2019-01 - 1	149TH-ST-E1-101	1381	0.2503	-	7471B	PPB	13 Apr 2021 14:58:46	U	SMPL
M2020-03 - 1	105-SB-1	4584	0.8541	-	7471B	PPB	13 Apr 2021 15:01:03	U	SMPL
M2017-01LX5 - 1		1122	0.2015	-	7471B	PPB	13 Apr 2021 15:03:19	U	SMPL
M2017-01A - 1		27441	5.1628	-	7471B	PPB	13 Apr 2021 15:05:35	U	SMPL
CCV95 - 1	CCV95	28151	5.2966	-	7471B	PPB	13 Apr 2021 15:07:52	U	SMPL
CCB95 - 1	CCB95	151	0.0185	-	7471B	PPB	13 Apr 2021 15:10:08	U	SMPL