


LB113961

245.1

INSTRUMENT ID: CV1

 μ Abs.:

40920



Linear

A= 0.0000e+000

B= 2.4460e-004

C= -1.1847e-002

Rho= 0.9997785

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9.1)
0.0	0.000	-0.018	-0.018	-27	0.000	-27					-
0.05	0.050	0.040	-0.010	210	0.0 %	210	0				-20
0.2	0.200	0.151	-0.049	666	0.0 %	666					-25
2.5	2.500	2.545	0.045	10452	0.0 %	10452					2
5.0	5.000	5.154	0.154	21121	0.0 %	21121					3
7.5	7.500	7.381	-0.119	30226	0.0 %	30226					-2
10.0	10.000	9.997	-0.003	40920	0.0 %	40920					0

LB113961 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Apr 2021 15:29:45

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-27	-	0.0000	245.1	PPB	13 Apr 2021 15:49:43	S	Std
0.05 - 1	50.05	210	-	0.0500	245.1	PPB	13 Apr 2021 15:54:35	S	Std
0.2 - 1	50.2	666	-	0.2000	245.1	PPB	13 Apr 2021 15:56:51	S	Std
2.5 - 1	52.5	10452	-	2.5000	245.1	PPB	13 Apr 2021 15:59:08	S	Std
5.0 - 1	55	21121	-	5.0000	245.1	PPB	13 Apr 2021 16:01:24	S	Std
7.5 - 1	57.5	30226	-	7.5000	245.1	PPB	13 Apr 2021 16:03:39	S	Std
10.0 - 1	51.8	40920	-	10.0000	245.1	PPB	13 Apr 2021 16:13:52	S	Std
ICV84 - 1	ICV84	17117	4.1750	-245.1		PPB	13 Apr 2021 16:19:11	U	SMPL
ICB84 - 1	ICB84	20	-0.0070	-245.1		PPB	13 Apr 2021 16:23:57	U	SMPL
CCV96 - 1	CCV96	21709	5.2982	-245.1		PPB	13 Apr 2021 16:26:12	U	SMPL
CCB96 - 1	CCB96	32	-0.0040	-245.1		PPB	13 Apr 2021 16:31:25	U	SMPL
CRA - 1	CRA	238	0.0464	-245.1		PPB	13 Apr 2021 16:36:32	U	SMPL
HighStd - 1	HighStd	40540	9.9043	-245.1		PPB	13 Apr 2021 16:38:47	U	SMPL
ChkStd - 1	ChkStd	28270	6.9030	-245.1		PPB	13 Apr 2021 16:41:02	U	SMPL
PB135709BL - 1	PBW	-33	-0.0199	-245.1		PPB	13 Apr 2021 16:43:20	U	SMPL
PB135709BS - 1	LCSW	14115	3.4407	-245.1		PPB	13 Apr 2021 16:45:36	U	SMPL
M1976-02 - 1	COMP	990	0.2303	-245.1		PPB	13 Apr 2021 16:47:53	U	SMPL
M1976-02DUP - 1	COMPDUP	915	0.2120	-245.1		PPB	13 Apr 2021 16:50:09	U	SMPL
M1976-02MS - 1	COMPMS	14807	3.6100	-245.1		PPB	13 Apr 2021 16:52:26	U	SMPL
M1976-02MSD - 1	COMPMSD	15055	3.6706	-245.1		PPB	13 Apr 2021 16:57:37	U	SMPL
M1990-01 - 1	ELEVATOR-PIT-STANDING-WATER	123	0.0182	-245.1		PPB	13 Apr 2021 16:59:52	U	SMPL
CCV97 - 1	CCV97	21551	5.2596	-245.1		PPB	13 Apr 2021 17:02:07	U	SMPL
CCB97 - 1	CCB97	-25	-0.0180	-245.1		PPB	13 Apr 2021 17:04:23	U	SMPL
M2007-01 - 1	55TH-ST-31	354	0.0747	-245.1		PPB	13 Apr 2021 17:06:38	U	SMPL
M2007-02 - 1	37TH-ST-2021	223	0.0427	-245.1		PPB	13 Apr 2021 17:08:53	U	SMPL
M1976-02LX5 - 1		303	0.0623	-245.1		PPB	13 Apr 2021 17:11:09	U	SMPL
M1976-02A - 1		15644	3.8147	-245.1		PPB	13 Apr 2021 17:13:24	U	SMPL
CCV98 - 1	CCV98	20804	5.0769	-245.1		PPB	13 Apr 2021 17:15:40	U	SMPL
CCB98 - 1	CCB98	-35	-0.0204	-245.1		PPB	13 Apr 2021 17:17:56	U	SMPL