



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113971

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/14/2021	11:00
2	CAL2	1	Water	NA	NA	20.1	7.00	04/14/2021	11:01
3	CAL3	1	Water	NA	NA	20.1	10.02	04/14/2021	11:02
4	ICV	1	Water	NA	NA	20.2	6.99	04/14/2021	11:05
5	CCV1	1	Water	NA	NA	20.3	2.01	04/14/2021	11:06
6	M2026-01	1	Water	NA	NA	20.3	6.63	04/14/2021	11:08
7	M2026-01DUP	1	Water	NA	NA	20.4	6.64	04/14/2021	11:09
8	M2026-02	1	Water	NA	NA	20.4	6.63	04/14/2021	11:10
9	M2026-03	1	Water	NA	NA	20.5	7.79	04/14/2021	11:11
10	M2026-04	1	Water	NA	NA	20.3	6.56	04/14/2021	11:15
11	M2026-05	1	Water	NA	NA	20.5	9.01	04/14/2021	11:18
12	M2026-06	1	Water	NA	NA	20.2	5.09	04/14/2021	11:21
13	M2038-01	1	Water	NA	NA	20.7	5.10	04/14/2021	11:24
14	M2039-01	1	Water	NA	NA	20.2	5.15	04/14/2021	11:25
15	CCV2	1	Water	NA	NA	20.3	12.02	04/14/2021	11:26

23 1397)

WORKLIST(Hardcopy Internal Chain)

WorkList Name: ph m2026 WorkList ID: 147993 Department: Wet-Chemistry Date: 04-14-2021 08:29:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/13/2021	Water	M2026-01	pH	Cool 4 deg C	WOOD06	O61	PZ-01	04/13/2021	9040C
04/13/2021	Water	M2026-02	pH	Cool 4 deg C	WOOD06	O61	PZ-02	04/13/2021	9040C
04/13/2021	Water	M2026-03	pH	Cool 4 deg C	WOOD06	O61	GMW-5	04/13/2021	9040C
04/13/2021	Water	M2026-04	pH	Cool 4 deg C	WOOD06	O61	PZ-15	04/13/2021	9040C
04/13/2021	Water	M2026-05	pH	Cool 4 deg C	WOOD06	O61	GMW-11	04/13/2021	9040C
04/13/2021	Water	M2026-06	pH	Cool 4 deg C	WOOD06	O61	FB-1	04/13/2021	9040C
04/09/2021	Water	M2038-01	pH	Cool 4 deg C	STER10	O62	POOL-WATER	04/09/2021	9040C
04/12/2021	Water	M2039-01	pH	Cool 4 deg C	STER11	O62	POOL-WATER	04/12/2021	9040C

Date/Time 04-14-21 10:50
Raw Sample Received by: JO (wet chem)
Raw Sample Relinquished by: AM (wet chem)

Date/Time 04-14-2021 12:10
Raw Sample Received by: AM (wet chem)
Raw Sample Relinquished by: JO (wet chem)