


7471B

LB113980
INSTRUMENT ID: CV1

Linear

A= 0.0000e+000

B= 1.9160e-004

C= -9.2925e-002

Rho= 0.9997631

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	901
0.0	0.000	0.067	0.067	836	0.000	836					-
0.2	0.200	0.202	0.002	1538	0.0 %	1538					1
2.5	2.500	2.461	-0.039	13332	0.0 %	13332					-2
5.0	5.000	4.998	-0.002	26572	0.0 %	26572					0
7.5	7.500	7.359	-0.141	38891	0.0 %	38891					-2
10.0	10.000	10.113	0.113	53265	0.0 %	53265					1

LB113980 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 14 Apr 2021 15:25:44					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	836	-	0.0000	7471B	PPB	14 Apr 2021 15:27:04	S	Std
0.2 - 1	50.2	1538	-	0.2000	7471B	PPB	14 Apr 2021 15:29:20	S	Std
2.5 - 1	52.5	13332	-	2.5000	7471B	PPB	14 Apr 2021 15:31:37	S	Std
5.0 - 1	55	26572	-	5.0000	7471B	PPB	14 Apr 2021 15:33:54	S	Std
7.5 - 1	57.5	38891	-	7.5000	7471B	PPB	14 Apr 2021 15:36:11	S	Std
10.0 - 1	510	53265	-	10.0000	7471B	PPB	14 Apr 2021 15:38:27	S	Std
ICV86 - 1	ICV86	20846	3.9012	-	7471B	PPB	14 Apr 2021 15:41:44	U	SMPL
ICB86 - 1	ICB86	354	-0.0251	-	7471B	PPB	14 Apr 2021 15:43:59	U	SMPL
CCV06 - 1	CCV06	27019	5.0839	-	7471B	PPB	14 Apr 2021 15:46:13	U	SMPL
CCB06 - 1	CCB06	348	-0.0262	-	7471B	PPB	14 Apr 2021 15:48:28	U	SMPL
CRA - 1	CRA	1561	0.2062	-	7471B	PPB	14 Apr 2021 15:50:43	U	SMPL
HighStd - 1	HighStd	53608	10.1784	-	7471B	PPB	14 Apr 2021 15:52:58	U	SMPL
ChkStd - 1	ChkStd	38028	7.1932	-	7471B	PPB	14 Apr 2021 15:55:13	U	SMPL
PB135731BL - 1	PBS	518	0.0063	-	7471B	PPB	14 Apr 2021 15:57:30	U	SMPL
PB135731BS - 1	LCSS	19205	3.5868	-	7471B	PPB	14 Apr 2021 15:59:48	U	SMPL
M2023-01 - 1	FK-GF-02-070-A	7694	1.3812	-	7471B	PPB	14 Apr 2021 16:07:00	U	SMPL
M2023-01DUP - 1	FK-GF-02-070-ADUP	7352	1.3157	-	7471B	PPB	14 Apr 2021 16:09:17	U	SMPL
M2023-01MS - 1	FK-GF-02-070-AMS	26095	4.9069	-	7471B	PPB	14 Apr 2021 16:11:34	U	SMPL
M2023-01MSD - 1	FK-GF-02-070-AMSD	26905	5.0621	-	7471B	PPB	14 Apr 2021 16:13:50	U	SMPL
M2023-03 - 1	FK-GF-02-070-B	5445	0.9503	-	7471B	PPB	14 Apr 2021 16:16:05	U	SMPL
CCV07 - 1	CCV07	27338	5.1450	-	7471B	PPB	14 Apr 2021 16:18:21	U	SMPL
CCB07 - 1	CCB07	377	-0.0207	-	7471B	PPB	14 Apr 2021 16:20:36	U	SMPL
M2024-01 - 1	FK-GF-02-071-A	9889	1.8018	-	7471B	PPB	14 Apr 2021 16:22:51	U	SMPL
M2024-03 - 1	FK-GF-02-071-B	6762	1.2027	-	7471B	PPB	14 Apr 2021 16:25:06	U	SMPL
M2025-04 - 1	104-SB-1	3164	0.5133	-	7471B	PPB	14 Apr 2021 16:27:22	U	SMPL
M2025-05 - 1	104-SB-2	3355	0.5499	-	7471B	PPB	14 Apr 2021 16:29:37	U	SMPL
M2025-07 - 1	WC-2	2114	0.3121	-	7471B	PPB	14 Apr 2021 16:31:52	U	SMPL
M2025-09 - 1	WC-3	4543	0.7775	-	7471B	PPB	14 Apr 2021 16:34:08	U	SMPL
M2025-11 - 1	130-SB-1	1557	0.2054	-	7471B	PPB	14 Apr 2021 16:36:25	U	SMPL
2023-01LX5 - 1		2008	0.2918	-	7471B	PPB	14 Apr 2021 16:38:42	U	SMPL
2023-01A - 1		29324	5.5256	-	7471B	PPB	14 Apr 2021 16:40:59	U	SMPL
CCV08 - 1	CCV08	27185	5.1157	-	7471B	PPB	14 Apr 2021 16:43:15	U	SMPL
CCB08 - 1	CCB08	378	-0.0205	-	7471B	PPB	14 Apr 2021 16:45:30	U	SMPL