



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113982

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/14/2021	17:15
2	CAL2	1	Water	NA	NA	20.3	7.01	04/14/2021	17:26
3	CAL3	1	Water	NA	NA	20.3	10.03	04/14/2021	17:27
4	ICV	1	Water	NA	NA	20.3	6.99	04/14/2021	17:30
5	CCV1	1	Water	NA	NA	20.3	2.01	04/14/2021	17:32
6	M2025-08	1	Solid	20.02	20	20.9	8.03	04/14/2021	17:36
7	M2025-10	1	Solid	20.01	20	21.1	6.41	04/14/2021	17:36
8	M2028-01	1	Solid	20.02	20	20.3	5.18	04/14/2021	17:38
9	M2028-03	1	Solid	20.03	20	22.8	6.32	04/14/2021	17:40
10	M2028-05	1	Solid	20.04	20	22.6	7.57	04/14/2021	17:41
11	M2028-07	1	Solid	20.03	20	22.7	6.19	04/14/2021	17:42
12	M2030-04	1	Solid	20.02	20	23.3	8.60	04/14/2021	17:33
13	M2030-04DUP	1	Solid	20.04	20	23.4	8.61	04/14/2021	17:34
14	M2046-03	1	Solid	20.04	20	22.2	7.16	04/14/2021	17:43
15	M2046-04	1	Solid	20.01	20	22.4	8.44	04/14/2021	17:44
16	CCV2	1	Water	NA	NA	20.5	12.02	04/14/2021	17:46

WORKLIST(Hardcopy Internal Chain)

113982

WorkList Name : corrosivity m2040

WorkList ID : 148015

Department : Wet-Chemistry

Date : 04-14-2021 17:18:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/13/2021	Solid	M2025-08	Corrosivity	Cool 4 deg C	DIST01	O62	WC-2	04/13/2021	9045D
04/13/2021	Solid	M2025-10	Corrosivity	Cool 4 deg C	DIST01	O62	WC-3	04/13/2021	9045D
04/14/2021	Solid	M2028-01	Corrosivity	Cool 4 deg C	PSEG03	O51	HC-1-OIL	04/14/2021	9045D
04/14/2021	Solid	M2028-03	Corrosivity	Cool 4 deg C	PSEG03	O51	HC-2-7-OILY-DEBRIS	04/14/2021	9045D
04/14/2021	Solid	M2028-05	Corrosivity	Cool 4 deg C	PSEG03	O51	HC-8	04/14/2021	9045D
04/14/2021	Solid	M2028-07	Corrosivity	Cool 4 deg C	PSEG03	O51	HC-9	04/14/2021	9045D
04/14/2021	Solid	M2030-04	Corrosivity	Cool 4 deg C	PSEG03	O51	ETGI-304	04/14/2021	9045D
04/14/2021	Solid	M2046-03	Corrosivity	Cool 4 deg C	DIST01	O51	WC-4	04/14/2021	9045D
04/14/2021	Solid	M2046-04	Corrosivity	Cool 4 deg C	DIST01	O51	WC-5	04/14/2021	9045D

Date/Time 04/14/21 17:20
 Raw Sample Received by: J.D. West (m)
 Raw Sample Relinquished by: J.D. West (m)

Date/Time 04/14/21 18:20
 Raw Sample Received by: J.C. (m)
 Raw Sample Relinquished by: J.D. West (m)