



## Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113983

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2718         |
| buffer solution pH 7 yellow    | W2627         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2809         |
| buffer solution pH 7 yellow    | W2719         |
| Buffer Solution, PH2 (500ml)   | W2806         |
| Buffer Solution, PH12 (500ml)  | W2807         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.2             | 4.01        | 04/14/2021 | 17:15     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.3             | 7.01        | 04/14/2021 | 17:26     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.3             | 10.03       | 04/14/2021 | 17:27     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.3             | 6.99        | 04/14/2021 | 17:30     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.3             | 2.01        | 04/14/2021 | 17:32     |
| 6   | M2030-03    | 1  | Solid  | 20.02       | 20          | 23.3             | 8.60        | 04/14/2021 | 17:33     |
| 7   | M2030-03DUP | 1  | Solid  | 20.04       | 20          | 23.4             | 8.61        | 04/14/2021 | 17:34     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.2             | 12.02       | 04/14/2021 | 17:35     |

# WORKLIST(Hardcopy Internal Chain)

28113983

WorkList Name : ph m2041

WorkList ID : 148016

Department : Wet-Chemistry

Date : 04-14-2021 17:19:18

| Due Date   | Matrix | Sample   | Test | Preservative | Customer | Raw Sample Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|------|--------------|----------|-----------------------------|-----------------|--------------|--------|
| 04/14/2021 | Solid  | M2030-03 | pH   | Cool 4 deg C | PSEG03   | O51                         | ETGI-304        | 04/14/2021   | 9045D  |

Date/Time 04-14-21 17:20  
 Raw Sample Received by: JD [Signature]  
 Raw Sample Relinquished by: [Signature]

Date/Time 04-14-21 18:20  
 Raw Sample Received by: J.C. (Sun)  
 Raw Sample Relinquished by: J.C. (Sun)