



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113984

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/14/2021	17:15
2	CAL2	1	Water	NA	NA	20.3	7.01	04/14/2021	17:26
3	CAL3	1	Water	NA	NA	20.3	10.03	04/14/2021	17:27
4	ICV	1	Water	NA	NA	20.3	6.99	04/14/2021	17:30
5	CCV1	1	Water	NA	NA	20.3	2.01	04/14/2021	17:32
6	M2033-01	1	Water	NA	NA	20.5	6.67	04/14/2021	17:47
7	M2033-01DUP	1	Water	NA	NA	20.6	6.68	04/14/2021	17:48
8	CCV2	1	Water	NA	NA	20.5	12.02	04/14/2021	17:49

AB 113984

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph w n2040 WorkList ID : 148017 Department : Wet-Chemistry Date : 04-14-2021 17:19:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/14/2021	Water	M2033-01	pH	Cool 4 deg C	PSEG03	O51	FM5210	04/14/2021	9040C

Date/Time 04/14/21 18:20
Raw Sample Received by: JN (wet ch(m))
Raw Sample Relinquished by: JN (wet ch(m))

Date/Time 04/14/21 18:20
Raw Sample Received by: JN (wet ch(m))
Raw Sample Relinquished by: JN (wet ch(m))