


7470A

LB 119987

INSTRUMENT ID: CV2

 μ Abs.:

39837



Linear

A= 0.0000e+000

B= 2.4751e-004

C= 6.2337e-003

Rho= 0.9995870

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/o 1)
0.0	0.000	0.037	0.037	124	0.000	124					
0.2	0.200	0.192	-0.008	749	0.0 %	749					-4
2.5	2.500	2.384	-0.116	9608	0.0 %	9608					-5
5.0	5.000	5.038	0.038	20331	0.0 %	20331					1
7.5	7.500	7.683	0.183	31014	0.0 %	31014					2
10.0	10.000	9.866	-0.134	39837	0.0 %	39837					-1

LB113987 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Apr 2021 12:32:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	124	-	0.0000	7470A	PPB	15 Apr 2021 12:35:07	S	Std
0.2 - 1	SD-2	749	-	0.2000	7470A	PPB	15 Apr 2021 12:37:24	S	Std
2.5 - 1	SD-5	9608	-	2.5000	7470A	PPB	15 Apr 2021 12:39:41	S	Std
5.0 - 1	SD	20331	-	5.0000	7470A	PPB	15 Apr 2021 12:41:57	S	Std
7.5 - 1	SD-5	31014	-	7.5000	7470A	PPB	15 Apr 2021 12:44:14	S	Std
10.0 - 1	SD	39837	-	10.0000	7470A	PPB	15 Apr 2021 12:49:08	S	Std
ICV87 - 1	ICV87	15727	3.8988	-7470A	PPB	15 Apr 2021 12:51:52	U	SMPL	
ICB87 - 1	ICB87	-74	-0.0121	-7470A	PPB	15 Apr 2021 12:54:07	U	SMPL	
CCV09 - 1	CCV09	20357	5.0448	-7470A	PPB	15 Apr 2021 12:56:21	U	SMPL	
CCB09 - 1	CCB09	-79	-0.0133	-7470A	PPB	15 Apr 2021 12:58:36	U	SMPL	
CRA - 1	CRA	736	0.1884	-7470A	PPB	15 Apr 2021 13:00:52	U	SMPL	
HighStd - 1	HighStd	40086	9.9279	-7470A	PPB	15 Apr 2021 13:03:07	U	SMPL	
ChkStd - 1	ChkStd	28774	7.1281	-7470A	PPB	15 Apr 2021 13:07:58	U	SMPL	
PB135749BL - 1	PBW	-105	-0.0198	-7470A	PPB	15 Apr 2021 13:10:13	U	SMPL	
PB135749BS - 1	LCSW	16428	4.0723	-7470A	PPB	15 Apr 2021 13:12:29	U	SMPL	
M2026-01 - 1	PZ-01	3714	0.9255	-7470A	PPB	15 Apr 2021 13:14:45	U	SMPL	
M2026-02 - 1	PZ-02	33125	8.2050	-7470A	PPB	15 Apr 2021 13:17:02	U	SMPL	
M2026-03 - 1	GMW-5	1014	0.2572	-7470A	PPB	15 Apr 2021 13:19:19	U	SMPL	
M2026-04 - 1	PZ-15	2135	0.5347	-7470A	PPB	15 Apr 2021 13:21:38	U	SMPL	
M2026-05 - 1	GMW-11	5159	1.2831	-7470A	PPB	15 Apr 2021 13:23:53	U	SMPL	
CCV10 - 1	CCV10	19483	4.8285	-7470A	PPB	15 Apr 2021 13:26:09	U	SMPL	
CCB10 - 1	CCB10	45	0.0174	-7470A	PPB	15 Apr 2021 13:28:24	U	SMPL	
M2026-06 - 1	FB-1	87	0.0278	-7470A	PPB	15 Apr 2021 13:30:40	U	SMPL	
M2026-10 - 1	PZ-01	92	0.0290	-7470A	PPB	15 Apr 2021 13:32:55	U	SMPL	
M2026-11 - 1	PZ-02	90	0.0285	-7470A	PPB	15 Apr 2021 13:35:10	U	SMPL	
M2026-12 - 1	GMW-5	99	0.0307	-7470A	PPB	15 Apr 2021 13:37:26	U	SMPL	
M2026-13 - 1	PZ-15	239	0.0654	-7470A	PPB	15 Apr 2021 13:39:42	U	SMPL	
M2026-14 - 1	GMW-11	144	0.0419	-7470A	PPB	15 Apr 2021 13:41:58	U	SMPL	
M2026-15 - 1	FB-1	108	0.0330	-7470A	PPB	15 Apr 2021 13:44:15	U	SMPL	
M2033-01 - 1	FM5210	338	0.0899	-7470A	PPB	15 Apr 2021 13:46:31	U	SMPL	
M2033-01DUP - 1	FM5210DUP	386	0.1018	-7470A	PPB	15 Apr 2021 13:48:48	U	SMPL	
M2033-01MS - 1	FM5210MS	13631	3.3800	-7470A	PPB	15 Apr 2021 13:51:03	U	SMPL	
CCV11 - 1	CCV11	19920	4.9366	-7470A	PPB	15 Apr 2021 13:53:19	U	SMPL	
CCB11 - 1	CCB11	30	0.0137	-7470A	PPB	15 Apr 2021 13:55:34	U	SMPL	
M2033-01MSD - 1	FM5210MSD	14929	3.7013	-7470A	PPB	15 Apr 2021 13:57:49	U	SMPL	
M2042-08 - 1	FS-SB12A	13431	3.3305	-7470A	PPB	15 Apr 2021 14:00:04	U	SMPL	
M2033-01LX5 - 1		82	0.0265	-7470A	PPB	15 Apr 2021 14:02:19	U	SMPL	
M2033-01A - 1		16254	4.0293	-7470A	PPB	15 Apr 2021 14:04:35	U	SMPL	
CCV12 - 1	CCV12	20035	4.9651	-7470A	PPB	15 Apr 2021 14:06:50	U	SMPL	
CCB12 - 1	CCB12	-26	-0.0002	-7470A	PPB	15 Apr 2021 14:09:07	U	SMPL	