

LB113990


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56023



A= 0.0000e+000

B= 1.7879e-004

C= -4.3139e-003

Rho= 0.9999794

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/0/0
0.0	0.000	0.020	0.020	135	0.000	135					-
0.2	0.200	0.210	0.010	1199	0.0 %	1199					5
2.5	2.500	2.453	-0.047	13744	0.0 %	13744					-2
5.0	5.000	5.019	0.019	28094	0.0 %	28094					0
7.5	7.500	7.487	-0.013	41899	0.0 %	41899					0
10.0	10.000	10.012	0.012	56023	0.0 %	56023					0

LB113990 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Apr 2021 14:43:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	135	-	0.0000	7471B	PPB	15 Apr 2021 14:47:54	S	Std
0.2 - 1	50.2	1199	-	0.2000	7471B	PPB	15 Apr 2021 14:50:10	S	Std
2.5 - 1	52.5	13744	-	2.5000	7471B	PPB	15 Apr 2021 14:52:28	S	Std
5.0 - 1	55	28094	-	5.0000	7471B	PPB	15 Apr 2021 14:54:44	S	Std
7.5 - 1	57.5	41899	-	7.5000	7471B	PPB	15 Apr 2021 14:57:01	S	Std
10.0 - 1	510	56023	-	10.0000	7471B	PPB	15 Apr 2021 14:59:17	S	Std
ICV88 - 1	ICV88	22098	3.9465	-	7471B	PPB	15 Apr 2021 15:01:57	U	SMPL
ICB88 - 1	ICB88	-14	-0.0068	-	7471B	PPB	15 Apr 2021 15:04:11	U	SMPL
CCV13 - 1	CCV13	28421	5.0770	-	7471B	PPB	15 Apr 2021 15:06:27	U	SMPL
CCB13 - 1	CCB13	-25	-0.0088	-	7471B	PPB	15 Apr 2021 15:08:42	U	SMPL
CRA - 1	CRA	1193	0.2090	-	7471B	PPB	15 Apr 2021 15:10:57	U	SMPL
HighStd - 1	HighStd	55702	9.9545	-	7471B	PPB	15 Apr 2021 15:13:12	U	SMPL
ChkStd - 1	ChkStd	38785	6.9300	-	7471B	PPB	15 Apr 2021 15:15:27	U	SMPL
PB135750BL - 1	PBS	-162	-0.0333	-	7471B	PPB	15 Apr 2021 15:17:45	U	SMPL
PB135750BS - 1	LCSS	19427	3.4690	-	7471B	PPB	15 Apr 2021 15:20:02	U	SMPL
M2029-01 - 1	OR-01-041421	2709	0.4800	-	7471B	PPB	15 Apr 2021 15:22:18	U	SMPL
M2034-03 - 1	1979	2850	0.5052	-	7471B	PPB	15 Apr 2021 15:24:35	U	SMPL
M2036-01 - 1	RT-80-W-ROCKAWAY-A1	2577	0.4564	-	7471B	PPB	15 Apr 2021 15:26:52	U	SMPL
M2036-03 - 1	RT-80-W-ROCKAWAY-A2	4459	0.7929	-	7471B	PPB	15 Apr 2021 15:29:09	U	SMPL
M2036-05 - 1	RT-80-W-ROCKAWAY-B1	6366	1.1338	-	7471B	PPB	15 Apr 2021 15:31:25	U	SMPL
CCV14 - 1	CCV14	28361	5.0663	-	7471B	PPB	15 Apr 2021 15:33:40	U	SMPL
CCB14 - 1	CCB14	11	-0.0023	-	7471B	PPB	15 Apr 2021 15:35:56	U	SMPL
M2036-07 - 1	RT-80-W-ROCKAWAY-B2	4255	0.7564	-	7471B	PPB	15 Apr 2021 15:38:11	U	SMPL
M2040-03 - 1	FILL-A	146743	26.2315	-	7471B	PPB	15 Apr 2021 15:40:26	U	SMPL
M2040-07 - 1	FILL-B	33455	5.9770	-	7471B	PPB	15 Apr 2021 15:42:41	U	SMPL
M2040-11 - 1	FILL-C	12826	2.2888	-	7471B	PPB	15 Apr 2021 15:45:00	U	SMPL
M2042-01 - 1	FS-SB12A(7-8)	9885	1.7630	-	7471B	PPB	15 Apr 2021 15:47:16	U	SMPL
M2042-01DUP - 1	FS-SB12A(7-8)DUP	8496	1.5147	-	7471B	PPB	15 Apr 2021 15:49:32	U	SMPL
M2042-02 - 1	M2042-01MS	29327	5.2390	-	7471B	PPB	15 Apr 2021 15:51:49	U	SMPL
M2042-03 - 1	M2042-01MSD	30237	5.4017	-	7471B	PPB	15 Apr 2021 15:54:05	U	SMPL
M2042-04 - 1	FS-SB12A(41-42)	319	0.0527	-	7471B	PPB	15 Apr 2021 15:56:22	U	SMPL
M2042-05 - 1	FS-SB12A(55-56)	187	0.0291	-	7471B	PPB	15 Apr 2021 15:58:38	U	SMPL
CCV15 - 1	CCV15	28795	5.1439	-	7471B	PPB	15 Apr 2021 16:00:53	U	SMPL
CCB15 - 1	CCB15	14	-0.0018	-	7471B	PPB	15 Apr 2021 16:03:09	U	SMPL
M2042-06 - 1	FS-SB12A(61-62)	207	0.0327	-	7471B	PPB	15 Apr 2021 16:05:24	U	SMPL
M2044-01 - 1	TR-05-041421	1310	0.2299	-	7471B	PPB	15 Apr 2021 16:07:40	U	SMPL
M2046-01 - 1	WC-4	446	0.0754	-	7471B	PPB	15 Apr 2021 16:09:55	U	SMPL
M2046-02 - 1	WC-5	2514	0.4452	-	7471B	PPB	15 Apr 2021 16:12:11	U	SMPL
M2047-05 - 1	130-SB3-(1.5-2)	28002	5.0021	-	7471B	PPB	15 Apr 2021 16:14:26	U	SMPL
M2042-01LX5 - 1		2006	0.3543	-	7471B	PPB	15 Apr 2021 16:16:42	U	SMPL
M2042-01A - 1		32075	5.7303	-	7471B	PPB	15 Apr 2021 16:18:59	U	SMPL
M2040-03DLX5 - 1		29732	5.3114	-	7471B	PPB	15 Apr 2021 16:21:15	U	SMPL
CCV16 - 1	CCV16	28737	5.1335	-	7471B	PPB	15 Apr 2021 16:23:33	U	SMPL
CCB16 - 1	CCB16	9	-0.0027	-	7471B	PPB	15 Apr 2021 16:25:49	U	SMPL