

LB114001


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4051E



A= 0.0000e+000

B= 2.4574e-004

C= -1.3071e-002

Rho= 0.9999336

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/05
0.0	0.000	0.013	0.013	108	0.000	108					
0.2	0.200	0.182	-0.018	795	0.0 %	795					-9
2.5	2.500	2.458	-0.042	10056	0.0 %	10056					-2
5.0	5.000	5.043	0.043	20574	0.0 %	20574					1
7.5	7.500	7.560	0.060	30816	0.0 %	30816					1
10.0	10.000	9.943	-0.057	40516	0.0 %	40516					-1

LB114001 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 16 Apr 2021 09:47:56					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	108	-	0.0000	7470A	PPB	16 Apr 2021 09:52:00	S	Std
0.2 - 1	50.2	795	-	0.2000	7470A	PPB	16 Apr 2021 09:54:17	S	Std
2.5 - 1	52.5	10056	-	2.5000	7470A	PPB	16 Apr 2021 09:56:35	S	Std
5.0 - 1	55	20574	-	5.0000	7470A	PPB	16 Apr 2021 09:58:51	S	Std
7.5 - 1	57.5	30816	-	7.5000	7470A	PPB	16 Apr 2021 10:01:08	S	Std
10.0 - 1	510	40516	-	10.0000	7470A	PPB	16 Apr 2021 10:05:57	S	Std
ICV89 - 1	ICV89	16965	4.1560	-	7470A	PPB	16 Apr 2021 10:08:38	U	SMPL
ICB89 - 1	ICB89	44	-0.0023	-	7470A	PPB	16 Apr 2021 10:13:34	U	SMPL
CCV17 - 1	CCV17	20297	4.9748	-	7470A	PPB	16 Apr 2021 10:15:49	U	SMPL
CCB17 - 1	CCB17	-135	-0.0462	-	7470A	PPB	16 Apr 2021 10:18:04	U	SMPL
CRA - 1	CRA	729	0.1661	-	7470A	PPB	16 Apr 2021 10:20:19	U	SMPL
HighStd - 1	HighStd	40724	9.9946	-	7470A	PPB	16 Apr 2021 10:22:33	U	SMPL
ChkStd - 1	ChkStd	28882	7.0845	-	7470A	PPB	16 Apr 2021 10:24:48	U	SMPL
PB135762BL - 1	PBW	-205	-0.0634	-	7470A	PPB	16 Apr 2021 10:27:05	U	SMPL
PB135762BS - 1	LCSW	15708	3.8471	-	7470A	PPB	16 Apr 2021 10:29:22	U	SMPL
M2025-08 - 1	WC-2	-141	-0.0477	-	7470A	PPB	16 Apr 2021 10:31:38	U	SMPL
M2025-08DUP - 1	WC-2DUP	17	-0.0089	-	7470A	PPB	16 Apr 2021 10:33:55	U	SMPL
M2025-08MS - 1	WC-2MS	17102	4.1896	-	7470A	PPB	16 Apr 2021 10:36:12	U	SMPL
M2025-08MSD - 1	WC-2MSD	17529	4.2946	-	7470A	PPB	16 Apr 2021 10:38:28	U	SMPL
M2025-10 - 1	WC-3	-69	-0.0300	-	7470A	PPB	16 Apr 2021 10:40:43	U	SMPL
CCV18 - 1	CCV18	20560	5.0394	-	7470A	PPB	16 Apr 2021 10:42:58	U	SMPL
CCB18 - 1	CCB18	-115	-0.0413	-	7470A	PPB	16 Apr 2021 10:45:13	U	SMPL
M2030-01 - 1	04-14-21-A	54	0.0002	-	7470A	PPB	16 Apr 2021 10:47:29	U	SMPL
M2030-04 - 1	ETGI-304	70	0.0041	-	7470A	PPB	16 Apr 2021 10:49:45	U	SMPL
M2034-04 - 1	1979	58	0.0012	-	7470A	PPB	16 Apr 2021 10:52:00	U	SMPL
M2043-01 - 1	FS-WC-SOIL	21	-0.0079	-	7470A	PPB	16 Apr 2021 10:54:16	U	SMPL
M2046-03 - 1	WC-4	141	0.0216	-	7470A	PPB	16 Apr 2021 10:56:32	U	SMPL
M2046-04 - 1	WC-5	95	0.0103	-	7470A	PPB	16 Apr 2021 10:58:47	U	SMPL
PB135763BL - 1	PBW	129	0.0186	-	7470A	PPB	16 Apr 2021 11:01:03	U	SMPL
PB135763BS - 1	LCSW	16946	4.1513	-	7470A	PPB	16 Apr 2021 11:03:20	U	SMPL
M2016-03 - 1	TOTE-1750	265	0.0521	-	7470A	PPB	16 Apr 2021 11:05:37	U	SMPL
M2030-02 - 1	04-14-21-WATER	38	-0.0037	-	7470A	PPB	16 Apr 2021 11:07:52	U	SMPL
CCV19 - 1	CCV19	21041	5.1576	-	7470A	PPB	16 Apr 2021 11:10:08	U	SMPL
CCB19 - 1	CCB19	-109	-0.0399	-	7470A	PPB	16 Apr 2021 11:12:23	U	SMPL
M2030-02DUP - 1	04-14-21-WATERDUP	39	-0.0035	-	7470A	PPB	16 Apr 2021 11:14:39	U	SMPL
M2030-02MS - 1	04-14-21-WATERMS	18490	4.5307	-	7470A	PPB	16 Apr 2021 11:16:55	U	SMPL
M2030-02MSD - 1	04-14-21-WATERMSD	15860	3.8844	-	7470A	PPB	16 Apr 2021 11:19:10	U	SMPL
M2032-02 - 1	HOLDING-TANK-260	-68	-0.0298	-	7470A	PPB	16 Apr 2021 11:21:26	U	SMPL
M2032-03 - 1	SECONDARY-TANK-262	70	0.0041	-	7470A	PPB	16 Apr 2021 11:23:42	U	SMPL
M2034-01 - 1	P1679	84	0.0076	-	7470A	PPB	16 Apr 2021 11:25:58	U	SMPL
CCV20 - 1	CCV20	20957	5.1370	-	7470A	PPB	16 Apr 2021 11:28:14	U	SMPL
CCB20 - 1	CCB20	-105	-0.0389	-	7470A	PPB	16 Apr 2021 11:30:30	U	SMPL
PB135762TB - 1	PB135725TB	48	-0.0013	-	7470A	PPB	16 Apr 2021 11:32:48	U	SMPL
PB135763TB - 1	PB135726TB	125	0.0176	-	7470A	PPB	16 Apr 2021 11:35:04	U	SMPL
M2025-08LX5 - 1		77	0.0059	-	7470A	PPB	16 Apr 2021 11:37:19	U	SMPL
M2025-08A - 1		16033	3.9269	-	7470A	PPB	16 Apr 2021 11:39:36	U	SMPL
M2030-02LX5 - 1		-63	-0.0286	-	7470A	PPB	16 Apr 2021 11:41:52	U	SMPL
M2030-02A - 1		16682	4.0864	-	7470A	PPB	16 Apr 2021 11:44:07	U	SMPL
CCV21 - 1	CCV21	19830	4.8600	-	7470A	PPB	16 Apr 2021 11:48:38	U	SMPL
CCB21 - 1	CCB21	18	-0.0086	-	7470A	PPB	16 Apr 2021 11:50:55	U	SMPL