



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114004

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	04/16/2021	09:45
2	M2028-01		1	12	100.00	>212.0	>212.0	04/16/2021	11:35
3	M2028-01DUP		1	12	100.00	>212.0	>212.0	04/16/2021	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp41621-1 WorkList ID : 148086 Department : Wet-Chemistry Date : 04-16-2021 09:19:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method	
04/20/2021	Solid	M2028-01	B	Flash Point	Cool 4 deg C	PSEG03	O51	HC-1-OIL	04/14/2021	1010A

Date/Time 04/16/21 09:25
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: KS (Wetchem)

Date/Time 04/16/21 12:15
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: KS (Wetchem)