


LB114007

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.9657e-004

C= -2.7923e-002

Rho= 0.9994361

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.0	0.000	-0.003	-0.003	128	0.000	128					-
0.2	0.200	0.165	-0.035	983	0.0 %	983					-18
2.5	2.500	2.414	-0.086	12421	0.0 %	12421					-3
5.0	5.000	5.133	0.133	26255	0.0 %	26255					3
7.5	7.500	7.677	0.177	39194	0.0 %	39194					2
10.0	10.000	9.814	-0.186	50068	0.0 %	50068					-2

LB114007 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Apr 2021 12:03:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	128	-	0.0000	7471B	PPB	16 Apr 2021 12:05:08	S	Std
0.2 - 1	50.2	983	-	0.2000	7471B	PPB	16 Apr 2021 12:07:24	S	Std
2.5 - 1	52.5	12421	-	2.5000	7471B	PPB	16 Apr 2021 12:09:42	S	Std
5.0 - 1	55	26255	-	5.0000	7471B	PPB	16 Apr 2021 12:11:59	S	Std
7.5 - 1	57.5	39194	-	7.5000	7471B	PPB	16 Apr 2021 12:14:15	S	Std
10.0 - 1	510	50068	-	10.0000	7471B	PPB	16 Apr 2021 12:21:47	S	Std
ICV90 - 1	ICV90	21540	4.2063	-	7471B	PPB	16 Apr 2021 12:24:25	U	SMPL
ICB90 - 1	ICB90	-43	-0.0364	-	7471B	PPB	16 Apr 2021 12:26:41	U	SMPL
CCV22 - 1	CCV22	27440	5.3660	-	7471B	PPB	16 Apr 2021 12:28:57	U	SMPL
CCB22 - 1	CCB22	-46	-0.0370	-	7471B	PPB	16 Apr 2021 12:31:12	U	SMPL
CRA - 1	CRA	1005	0.1696	-	7471B	PPB	16 Apr 2021 12:33:28	U	SMPL
HighStd - 1	HighStd	51094	10.0158	-	7471B	PPB	16 Apr 2021 12:35:42	U	SMPL
ChkStd - 1	ChkStd	37946	7.4312	-	7471B	PPB	16 Apr 2021 12:37:57	U	SMPL
PB135769BL - 1	PBS	-54	-0.0385	-	7471B	PPB	16 Apr 2021 12:40:15	U	SMPL
PB135769BS - 1	LCSS	18963	3.6997	-	7471B	PPB	16 Apr 2021 12:42:32	U	SMPL
M2048-01 - 1	SB-01-0-2.0	13763	2.6775	-	7471B	PPB	16 Apr 2021 12:44:48	U	SMPL
M2048-03 - 1	SB-01-14.5-15.0	744	0.1183	-	7471B	PPB	16 Apr 2021 12:47:04	U	SMPL
M2048-04 - 1	SB-02-0-2.0	21925	4.2819	-	7471B	PPB	16 Apr 2021 12:49:21	U	SMPL
M2048-06 - 1	SB-02-13.5-14.0	209	0.0132	-	7471B	PPB	16 Apr 2021 12:51:38	U	SMPL
M2048-07 - 1	SB-03-0-2.0	7676	1.4810	-	7471B	PPB	16 Apr 2021 12:53:53	U	SMPL
CCV23 - 1	CCV23	27404	5.3590	-	7471B	PPB	16 Apr 2021 12:56:09	U	SMPL
CCB23 - 1	CCB23	-36	-0.0350	-	7471B	PPB	16 Apr 2021 12:58:24	U	SMPL
M2048-09 - 1	SB-03-14.5-15.0	1462	0.2595	-	7471B	PPB	16 Apr 2021 13:00:40	U	SMPL
M2048-10 - 1	SB-04-0-2.0	5311	1.0161	-	7471B	PPB	16 Apr 2021 13:02:56	U	SMPL
M2048-12 - 1	SB-04-14.5-15.0	4615	0.8793	-	7471B	PPB	16 Apr 2021 13:05:11	U	SMPL
M2048-13 - 1	SB-05-0-2.0	1020	0.1726	-	7471B	PPB	16 Apr 2021 13:07:26	U	SMPL
M2048-15 - 1	SB-05-14.5-15.0	248	0.0208	-	7471B	PPB	16 Apr 2021 13:09:41	U	SMPL
M2048-16 - 1	SB-06-0-2.0	1224	0.2127	-	7471B	PPB	16 Apr 2021 13:11:58	U	SMPL
M2048-18 - 1	SB-06-8.5-9.0	469	0.0643	-	7471B	PPB	16 Apr 2021 13:14:14	U	SMPL
M2048-19 - 1	SB-07-0-2.0	457	0.0619	-	7471B	PPB	16 Apr 2021 13:21:29	U	SMPL
M2048-19DUP - 1	SB-07-0-2.0DUP	512	0.0727	-	7471B	PPB	16 Apr 2021 13:23:46	U	SMPL
M2048-19MS - 1	SB-07-0-2.0MS	21713	4.2403	-	7471B	PPB	16 Apr 2021 13:26:02	U	SMPL
CCV24 - 1	CCV24	27649	5.4071	-	7471B	PPB	16 Apr 2021 13:28:17	U	SMPL
CCB24 - 1	CCB24	-17	-0.0313	-	7471B	PPB	16 Apr 2021 13:30:33	U	SMPL
M2048-19MSD - 1	SB-07-0-2.0MSD	19948	3.8933	-	7471B	PPB	16 Apr 2021 13:32:48	U	SMPL
PB135770BL - 1	PBS	39	-0.0203	-	7471B	PPB	16 Apr 2021 13:35:03	U	SMPL
PB135770BS - 1	LCSS	18810	3.6696	-	7471B	PPB	16 Apr 2021 13:37:19	U	SMPL
M2049-01 - 1	SB-07-9.5-10.0	585	0.0871	-	7471B	PPB	16 Apr 2021 13:39:34	U	SMPL
M2049-01DUP - 1	SB-07-9.5-10.0DUP	577	0.0855	-	7471B	PPB	16 Apr 2021 13:41:49	U	SMPL
M2049-01MS - 1	SB-07-9.5-10.0MS	20348	3.9719	-	7471B	PPB	16 Apr 2021 13:44:05	U	SMPL
M2049-01MSD - 1	SB-07-9.5-10.0MSD	20915	4.0834	-	7471B	PPB	16 Apr 2021 13:46:22	U	SMPL
M2049-02 - 1	SB-08-0-2.0	549	0.0800	-	7471B	PPB	16 Apr 2021 13:48:39	U	SMPL
M2049-04 - 1	SB-08-7.5-8.0	769	0.1232	-	7471B	PPB	16 Apr 2021 13:50:56	U	SMPL
M2049-05 - 1	SB-08-7.5-8.0-DUP	302	0.0314	-	7471B	PPB	16 Apr 2021 13:53:12	U	SMPL
CCV25 - 1	CCV25	27811	5.4390	-	7471B	PPB	16 Apr 2021 13:55:27	U	SMPL
CCB25 - 1	CCB25	23	-0.0234	-	7471B	PPB	16 Apr 2021 13:57:43	U	SMPL
M2049-06 - 1	SB-09-0-2.0	8669	1.6762	-	7471B	PPB	16 Apr 2021 13:59:58	U	SMPL
M2049-08 - 1	SB-09-5.5-6.0	441	0.0588	-	7471B	PPB	16 Apr 2021 14:02:13	U	SMPL
M2049-10 - 1	SB-10-0-2.0	1683	0.3029	-	7471B	PPB	16 Apr 2021 14:04:29	U	SMPL
M2049-12 - 1	SB-10-8.5-9.0	256	0.0224	-	7471B	PPB	16 Apr 2021 14:06:44	U	SMPL
M2049-13 - 1	SB-11-0-2.0	2006	0.3664	-	7471B	PPB	16 Apr 2021 14:09:00	U	SMPL
M2049-15 - 1	SB-11-8.5-9.0	257	0.0226	-	7471B	PPB	16 Apr 2021 14:11:16	U	SMPL
M2049-16 - 1	SB-12-0-2.0	1933	0.3521	-	7471B	PPB	16 Apr 2021 14:13:32	U	SMPL
M2049-18 - 1	SB-12-13.5-14.0	264	0.0240	-	7471B	PPB	16 Apr 2021 14:15:49	U	SMPL
CCV26 - 1	CCV26	27885	5.4535	-	7471B	PPB	16 Apr 2021 14:18:06	U	SMPL
CCB26 - 1	CCB26	41	-0.0199	-	7471B	PPB	16 Apr 2021 14:20:21	U	SMPL
M2048-19LX5 - 1		165	0.0045	-	7471B	PPB	16 Apr 2021 14:22:36	U	SMPL
M2048-19A - 1		22172	4.3305	-	7471B	PPB	16 Apr 2021 14:24:52	U	SMPL
M2049-01LX5 - 1		89	-0.0104	-	7471B	PPB	16 Apr 2021 14:27:07	U	SMPL
M2049-01A - 1		22331	4.3617	-	7471B	PPB	16 Apr 2021 14:29:23	U	SMPL
CCV27 - 1	CCV27	26104	5.1034	-	7471B	PPB	16 Apr 2021 14:34:20	U	SMPL
CCB27 - 1	CCB27	1	-0.0277	-	7471B	PPB	16 Apr 2021 14:36:35	U	SMPL