



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114009

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	04/16/2021	13:02
2	CAL2	1	Water	NA	NA	21.3	7.00	04/16/2021	13:03
3	CAL3	1	Water	NA	NA	21.3	10.02	04/16/2021	13:04
4	ICV	1	Water	NA	NA	21.3	6.99	04/16/2021	13:05
5	CCV1	1	Water	NA	NA	21.3	2.01	04/16/2021	13:06
6	M2043-01	1	Solid	20.03	20	21.4	10.01	04/16/2021	13:10
7	M2043-01DUP	1	Solid	20.04	20	21.5	10.02	04/16/2021	13:11
8	M2053-01	1	Solid	20.02	20	20.8	8.13	04/16/2021	13:12
9	CCV2	1	Water	NA	NA	20.2	12.02	04/16/2021	13:15

WORKLIST(Hardcopy Internal Chain)

28114009

WorkList Name : PH M2043 W WorkList ID : 148094 Department : Wet-Chemistry Date : 04-16-2021 12:54:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/14/2021	Solid	M2043-01	pH	Cool 4 deg C	URSC01	O51	FS-WC-SOIL	04/14/2021	9045D
04/15/2021	Solid	M2053-01	pH	Cool 4 deg C	PSEG03	O62	PADS-2	04/15/2021	9045D

Date/Time 04-16-21 13:00
 Raw Sample Received by: JN West Chem,
 Raw Sample Relinquished by: JN West Chem

Date/Time 04-16-21 14:35
 Raw Sample Received by: JN West Chem
 Raw Sample Relinquished by: JN West Chem