



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114011

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	04/16/2021	13:02
2	CAL2	1	Water	NA	NA	21.3	7.00	04/16/2021	13:03
3	CAL3	1	Water	NA	NA	21.3	10.02	04/16/2021	13:04
4	ICV	1	Water	NA	NA	21.3	6.99	04/16/2021	13:05
5	CCV1	1	Water	NA	NA	21.3	2.01	04/16/2021	13:06
6	M2073-01	1	Water	NA	NA	20.6	5.41	04/16/2021	13:20
7	M2073-01DUP	1	Water	NA	NA	20.7	5.42	04/16/2021	13:21
8	CCV2	1	Water	NA	NA	20.8	12.02	04/16/2021	13:22

18114011

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2073 WorkList ID : 148095 Department : Wet-Chemistry Date : 04-16-2021 12:55:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/15/2021	Water	M2073-01	pH	Cool 4 deg C	STER01	P11	POOL-WATER	04/15/2021	SM 4500-H B

Date/Time 04.16.21 13:00
 Raw Sample Received by: JO [Signature]
 Raw Sample Relinquished by: JO [Signature]

Date/Time 04.16.21 14:35
 Raw Sample Received by: JO [Signature]
 Raw Sample Relinquished by: JO [Signature]